

seems OK- 14 ppm to be checked-Request to provide Input on Published Amendments to Air Quality Standards

AV avkumar@barc.gov.in
Wed, 11 Sep 2024 3:59:23 AM +0530 •

To "ChdTCThirtyFive Chemical Engineering" <chd35@bis.org.in>

Cc "ggp369" <ggp369@yahoo.co.in>, "ggp358" <ggp358@gmail.com>, "sksahu" <sksahu@barc.gov.in>, "shankar" <aggarwalsg@nplindia.org>, "aggarwalshan" <aggarwalshan@gmail.com>, "Tuhin Kumar Mandal" <tuhin@nplindia.org>, "Sanjeev Goyal" <sk_goyal@neeri.res.in>, "aggarwalsmita8" <aggarwalsmita8@gmail.com>, "Pawan Labhasetwar" <pk_labhasetwar@neeri.res.in>, "dg.cpcb" <dg.cpcb@gmail.com>, "rajeshcs cpcb" <rajeshcs.cpcb@nic.in>, "adiman25sharma" <adiman25sharma@gmail.com>, "mohit.sharma" <mohit.sharma@cii.in>, "shubham.mishra" <shubham.mishra@cii.in>, "shekh.faridi" <shekh.faridi@dyson.com>, "ecotechinstruments" <ecotechinstruments@gmail.com>, "eeplkol" <eeplkol@gmail.com>, "eeplkol2" <eeplkol2@gmail.com>, "sanjib36" <sanjib36@yahoo.com>, "sbalbir058" <sbalbir058@gmail.com>, "ashishgupta.envirotech" <ashishgupta.envirotech@gmail.com>, "greeneconomyipl" <greeneconomyipl@gmail.com>, "reggreeneconomyipl" <reggreeneconomyipl@gmail.com>, "sharmajswarup" <sharmajswarup@hotmail.com>, "iaapcdelhi" <iaapcdelhi@gmail.com>, "dhrumilsoni" <dhrumilsoni@iccmil.in>, "regulatory" <regulatory@iccmil.in>, "iccnhq" <iccnhq@iccmil.in>, "iccmumbai" <iccmumbai@iccmil.in>, "kishorgawankar" <kishorgawankar@mpcb.gov.in>, "padmanabh.khadkikar" <padmanabh.khadkikar@mpcb.gov.in>, "rajivranjan02" <rajivranjan02@ntpc.co.in>, "rajeshmalik" <rajeshmalik@ntpc.co.in>, "Anand Bohra" <anand.bohra@ncbindia.com>, "cme.env" <cme.env@gmail.com>, "K R Pundareeka Nath" <krp.nath@ncbindia.com>, "pundareekanath" <pundareekanath@gmail.com>, "tech" <tech@faidelhi.org>, "arun.mondal" <arun.mondal@indorama.com>, "naikrc" <naikrc@uniphos.com>, "vijay.pandey" <vijay.pandey@uniphos-envirotronic.com>, "vinayak.valsangkar" <vinayak.valsangkar@uniphos-envirotronic.com>, "nraveendhar" <nraveendhar@rediffmail.com>, "cpcbskt" <cpcbskt@yahoo.co.in>, "rizvi.irdh" <rizvi.irdh@gmail.com>, "nrdrizvi" <nrdrizvi@gmail.com>

Tags

No additional comments.

Note : w.r.t. to the comment received from Mr. Vinayak Valsangkar, "Page 3, 9.2 Para 1 Inserting a line after 60 mg/m³: Need 14 ppm in place of 4ppm"

How 14 ppm is obtained needs to be re-checked.

What I could understand, from page 3, section 9.2 (IS 15206:2002/ISO 8760:1990), is :

1. Maximum error allowed is 25% of 60 mg/m³. This is equal to $0.25 \times 60 = 15$ mg/m³.
2. In page 3, section 9.1, last line, it is given that 1 mg/m³ = 0.86 ppm i.e. 15 mg/m³ will be equal to $15 \times 0.86 = 12.9$ ppm.

So it should be 12.9 ppm or say 13 ppm, NOT 14 ppm.

सादर / With Regards,

डॉ. ए. विनोद कुमार/ Dr. A. Vinod Kumar

उत्कृष्ट वैज्ञानिक एवं अध्यक्ष, पर्यावरणीय अनुवीक्षण एवं मूल्यांकन प्रभाग/

Outstanding Scientist & Head, Environmental Monitoring and Assessment Division,

भाभा परमाणु अनुसंधान केंद्र / Bhabha Atomic Research Centre,

ट्रॉम्बे, मुंबई / Trombay, Mumbai - 400 085

दूरभाष / Phone: 22-25595380/69295380

फैक्स/Fax : 022-25505313/69295313

वरिष्ठ प्रोफेसर, होमी भाभा राष्ट्रीय संस्थान, अणुशक्तिनगर, मुंबई/Senior

Professor, Homi Bhabha National Institute, Anushaktinagar, Mumbai 400094

On 2024-09-06 11:43, ChdTCThirtyFive Chemical Engineering wrote:

> Dear Members of CHD 35,

>

> This is with reference to the 18th Meeting of CHD 35, Air Quality
> Sectional Committee held on 13-06-2024. During the discussion in the
> meeting , Committee decided that the amendment published on the below
> mentioned standards along with the comments received from Mr. Vinayak
> Valsangkar will be circulated to the members for their Inputs.

>

> * IS 15206:2002/ISO 8760:1990 - Work - Place air - Determination of
> mass concentration of carbon monoxide - Method using detector tubes
> for short-term sampling with direct indication

>

> * IS 15209 : 2002 /ISO 8761 : 1989 - Work - Place air - Determination
> of mass concentration of nitrogen dioxide - Method using detector
> tubes for short term sampling with direct indication

>

> Your valuable insights and feedback on whether these amendments should
> be adopted by India or not are highly appreciated. Kindly review the
> attached documents and provide your inputs preferably before

>

> 15-09-2024. So that the inputs received from the members can be
> discussed in the main committee meeting on 19-09-2024 .

>

> Best Regards,

>

> Preeti Prabha

> Scientist 'C' | Chemical Department | Member Secretary

> Air Quality Sectional Committee, CHD 35

- > BUREAU OF INDIAN STANDARDS
 - > Manak Bhawan, 9-Bahadur Shah Zafar Marg, New Delhi - 110002, INDIA
 - > chd35@bis.org.in
 - > ☎ 011 23608532 |
 - > Learn more about BIS: www.bis.gov.in
-