

भारतीय मानक  
Indian Standard

IS 14194 (Part 3/Sec 1) : 2024

---

---

पर्यावरण नमूनों में रेडियोन्यूक्लाइड —  
आंकलन की पद्धतियाँ

भाग 3 यूरेनियम  
अनुभाग 1 पानी के नमूने में  
( दूसरा पुनरीक्षण )

**Radionuclides in Environmental  
Samples — Methods of Estimation**  
**Part 3 Uranium**  
**Section 1 In Water Sample**  
( *Second Revision* )

ICS 13.020.40; 13.280

© BIS 2024



भारतीय मानक ब्यूरो  
BUREAU OF INDIAN STANDARDS  
मानक भवन, 9 बहादुर शाह ज़फर मार्ग, नई दिल्ली - 110002  
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG  
NEW DELHI - 110002  
[www.bis.gov.in](http://www.bis.gov.in) [www.standardsbis.in](http://www.standardsbis.in)

March 2024

Price Group 6

## FOREWORD

This Indian Standard (Part 3/Sec 1) (Second Revision) was adopted by the Bureau of Indian Standards, after the draft finalized by the Nuclear Energy for Peaceful Applications Sectional Committee, had been approved by the Chemical Division Council.

Radioactivity from several naturally occurring and anthropogenic sources is present throughout the environment. Uranium, radium, strontium ( $^{90}\text{Sr}$ ), iodine ( $^{131}\text{I}$ ) and caesium ( $^{137}\text{Cs}$ ) are the radionuclides of importance in environmental radioactivity monitoring. Uranium occurs in natural waters as a composite of three isotopes ( $^{234}\text{U}$ ,  $^{235}\text{U}$  and  $^{238}\text{U}$ ) at low specific activity of ppb ( $\mu\text{g/l}$ ) or sub-ppb levels. Except uranium, other radionuclides in environmental samples are evaluated by separation and the radioactivity is measured by either of alpha, beta or gamma radiation, emitted by the radionuclides.

This standard was first published in 1994 for measurement of uranium by fluorimetric method, which is very sensitive compared to radiation measurement. In the first revision in 2021, terminology given in the standard was modified to align them with the terminology given in other parts of IS 14194. Alternate method for preparation of water sample was also prescribed.

While reviewing this standard, the committee has decided to further divide this standard into two sections, one for only water sample and other for geological and biological samples for the purpose of better understanding and clarity. This section covers the measurement of uranium in water sample (for example, surface water, ground water and sea water). IS 14194 (Part 3/Sec 2) covers uranium measurement in geological and biological samples like soil, sediment, rock, vegetation, food and biota samples.

All laboratories shall be duly certified by Atomic Energy Regulatory Board, Government of India for carrying out analysis or measurement of radioactivity in commodities.

The composition of the Committee responsible for the formulation of this standard is given in Annex B.

In reporting the result of a test or analysis made in accordance with this standard, if the final value, observed or calculated, is to be rounded off, it shall be done in accordance with IS 2 : 2022 'Rules for rounding off numerical values (*second revision*)'.

*Indian Standard***RADIONUCLIDES IN ENVIRONMENTAL SAMPLES —  
METHODS OF ESTIMATION****PART 3 URANIUM****SECTION 1 IN WATER SAMPLE***( Second Revision )***1 SCOPE**

The method described in this standard uses light-emitting diode (LED) fluorimetry to estimate the total uranium content of water sample.

**2 REFERENCES**

The standard given below contain provisions which, through reference in this text, constitute provisions of this standard. At the time of publication, the edition indicated was valid. All standards are subject to revisions, and parties to agreements based on this standard are encouraged to investigate the possibility of applying the most recent edition of this standard:

<i>IS No.</i>	<i>Title</i>
IS 1070 : 2023	Reagent grade water — Specification ( <i>fourth revision</i> )

**3 TERMINOLOGY****3.1 Activity**

The number of spontaneous nuclear disintegrations occurring in a certain quantity (weight/volume) of material during a specified time interval, divided by that time interval. It is expressed in Becquerel (Bq), where 1 Bq = 1 disintegration/ second. Earlier unit of activity was Curie (Ci), where 1 Ci =  $3.7 \times 10^{10}$  disintegrations/second.

**3.2 Radionuclide**

An element, with a certain mass number and atomic number, which undergoes spontaneous disintegration in a measurable period of time.

**4 PRINCIPLE**

Uranium present in a sample forms a complex with sodium-pyrophosphate, a fluorescence enhancing agent, in a 9 ml suprasil quartz cuvette. This complex is excited by LED at an excitation wavelength of 410 nm from an array of LEDs (7 numbers). The uranium complex, then fluoresces with emission peaks at 494 nm, 516 nm, 540 nm and 565 nm, with maximum at ~ 510 nm. The fluorescent radiation is detected by a photomultiplier tube (PMT). The detector is present at right angle to

the radiation source, thus the background signal due to the source will be negligible and a lower detection level can be achieved than atomic absorption or emission techniques.

The fluorescence of organic matters excited by the nitrogen laser have very short lifetime ( $< 100$  ns) compared to that of uranyl complex ( $\sim 200$   $\mu$ s). By measuring the delayed fluorescence signal (only for a 100  $\mu$ s period starting from 30  $\mu$ s after the laser pulse) and fluorescence signals from the PMT are integrated for 4 s and then displayed. Thus, the fluorescence due to organic compounds dies down and only the fluorescence of uranium isotopes collected by the time gated PMT. The fluorescence is proportional to concentration of uranium in aqueous medium.

Otherwise, the fluorescence peak of organic matters is around 400 nm and that of uranium in complexing media are 494 nm, 516 nm, 540 nm and 565 nm, maximum around 510 nm. Therefore, using an optical filter at 450 nm, the contribution of organic matters is nullified. Thus, the fluorescence interference due to organic matters is corrected by using proper optical filter, fixed electronic delay and gating technology. The quenching effect encountered due to the presence of high concentration of acid (up to 2 M) during analysis, is removed by a proper pH adjustment prior to analysis.

**5 APPARATUS****5.1 LED Uranium Fluorimeter**

The total uranium content in the water samples is analysed using a uranium fluorimeter system. It is a compact analytical instrument based on fluorimetric technique and micro-optics. The instrument contains LED/laser source, sample compartment and a PMT placed in right-angle optical configuration. The fluorescence interference originating from organic matters and other metals in the samples, is corrected by optical filter, fixed electronic delay and gating. Uranium concentration in the samples can be estimated using the standard addition method, to avoid matrix effect. The working range and minimum detection level (MDL)

of the instrument shall be 0.2 µg/l to 20 µg/l and 0.2 µg/l, respectively.

**5.2 Micro-Pipettes** — (5 ml, 0.5 ml, 100 µl, 50 µl) with disposable tips.

**5.3 Analytical Weighting Balance**

**5.4 pH Meter**

**5.5 TDS Meter**

**5.6 Suprasil Quartz Cuvette with Teflon Cap**

**5.7 Lint Free Wipers**

**5.8 Uranium Standard Solution (100 ppm or Higher)**

**5.9 Sodium Pyrophosphate** — purity more than 99 percent.

**5.10 Phosphoric Acid** — purity more than 90 percent.

## 6 REAGENTS

### 6.1 Stock Solution

A stock solution of 0.973 g/l uranium (Aldrich make) is diluted to working concentration using ultrapure water (*see* Note) for regular calibration of the system.

NOTE — Ultrapure water shall have specification given in Table 1.

**Table 1 Requirement of Ultrapure water**

(Clause 6.1)

Sl No. (1)	Parameter (2)	Requirement (3)
i)	Resistivity (25 °C)	> 18.18 MOhm
ii)	Total organic carbon (on-line for < 10 ppb)	< 1 µg/l
iii)	On-line dissolved oxygen	10 µg/l

**6.2 Fluorescence Enhancing Agent** — 5 percent sodium pyrophosphate (SPP).

Dissolve 1.25 g of sodium pyrophosphate (SPP) in 20 ml of ultrapure water and add drop-wise phosphoric acid to adjust pH ~ 7 and then make up the volume up to 25 ml using ultrapure water. Label the date of preparation and final pH of the solution on the standard flask.

NOTE — All the volumetric additions to be carried out using both micro-pipettes and analytical weighing balance to confirm amount by volumetry and gravimetry.

## 7 MEASUREMENT PROCEDURE

**7.1** Switch on the instrument and allow it to warm-up and equilibrate with room temperature for one hour.

Calibrate the fluorimeter with four uranium standards of increasing concentrations to check the instrument performance and the linear dynamic range. For zero concentration, pipette out 5 ml of double distilled water onto a dry suprasil quartz cuvette and add 0.5 ml of buffer that is 5 percent SPP, mix well and record the system blank ( $B \pm \sigma$ ). Fig. 1 shows a typical calibration graph for fluorimeter response using standard uranium solutions. The water samples have to be filtered

using 0.45 micron filter papers and a vacuum filtration assembly before analysis to remove the suspended solids. If the total dissolved solids (TDS) level is low (< 1 500 ppm), then the water sample can be directly analysed for uranium using fluorimeter, no chemical processing is required. The sample and standard are maintained at room temperature before analysis by the equipment in the same room. High TDS samples, like seawater, are treated in a separate manner, the details of processing of seawater samples are given in Annex A.

**7.2** Pipette out 5 ml of water sample onto a cleaned and dry cuvette, add 0.5 ml of 5 percent SPP (buffer with pH ~ 7). Record the fluorescence response of the sample only, in terms of counts, minimum 4 readings.

**7.3** Add 50 µl of 500 µg/l uranium standard, gravimetrically and volumetrically, onto the cuvette that contains the sample and buffer, record the fluorescence response after the first standard addition.

**7.4** Repeat the above sequence, that is again add 50 µl of 500 µg/l uranium standard onto the cuvette and record the fluorescence response after second standard addition.




FIG. 1 CALIBRATION GRAPH FOR FLUORIMETER RESPONSE USING STANDARD URANIUM SOLUTION

**7.5** The concentration of uranium (ppb or  $\mu\text{g/l}$ ) in samples was calculated by using the formula,

Uranium concentration (ppb or  $\mu\text{g/l}$ ) =

$$\frac{D_1}{(D_2 - D_1)} \times \frac{(V_2 \times C)}{V_1}$$

where

$D_1$  = fluorescence due to sample only (blank subtracted);

$D_2$  = fluorescence due to sample and uranium standard (blank subtracted);

$V_2$  = volume of uranium standard (ml);

$V_1$  = volume of sample taken (ml); and

$C$  = concentration of uranium standard (ppb or  $\mu\text{g/l}$ ).

In case of multiple standard additions, the counts from 7.2, 7.3, and 7.4 steps above are populated in a table and then fluorescence data is plotted against concentration of uranium present in the sample after standard addition. From the slope of the graph, counts per  $\mu\text{g/l}$  are obtained. Dividing the net counts for sample (blank subtracted sample fluorescence) with the value of slope, uranium concentration in sample is calculated.

### 7.6 Error Propagation

Fluorescence count for the sample  $S = S \pm \sigma_S$

Blank count for the instrument  $B = B \pm \sigma_B$

Net count =  $D \pm \sigma_D = (S - B) \pm (\sigma_S^2 + \sigma_B^2)$

where

$$D = S - B; \text{ and}$$

$$\sigma_D = (\sigma_S^2 + \sigma_B^2).$$

## 8 CALCULATION

### 8.1 In Absence of Matrix Effect (No Quenching)

This method is applicable for synthetic samples where concentrations of quench inducing elements are zero or negligible. To estimate uranium concentration in such samples, evaluate the calibration factor  $F$  (ppb/count or  $\mu\text{g/l}$  per count of uranium) from the slope of calibration curve.

Uranium concentration in ppb or  $\mu\text{g/l}$  in sample =

$$(S - B) \times F$$

where

$S$  = fluorescence count for the sample; and

$B$  = blank count for the instrument (only 5 ml distilled water and 0.5 ml SPP).

### 8.2 In Presence of Matrix Effect

Most of water samples are of natural origin, for example groundwater, surface water etc which contains different levels of TDS concentrations and may have unknown amount of elements contributing to quenching of uranium fluorescence. In this case standard addition method has to be followed to correct the matrix effect. Quenching due to matrix effect can be detected from sample counts and sample plus standard counts. If the latter is not nearly equal to the sum of counts of sample and standard measured separately, quenching is present. In such cases, standard addition method is followed.

Standard addition method involves analysing solutions containing the same amount of the sample to be analysed, but increasing amounts of standard being added. The uranium concentration in the sample is unknown but in the standard is known.

For example, 5 ml of unknown sample is first measured, then 0.05 ml of uranium standard is added

to it before two successive measurements. The idea of this procedure is that the total uranium concentration is the sum of the uranium present in the unknown sample and known standard, and that

the total uranium concentration varies linearly. If the signal response is linear in this concentration range, then a plot similar to Fig. 2 is generated.




FIG. 2 STANDARD ADDITION GRAPH FOR UNKNOWN SAMPLE

### 9 QA/QC

**9.1** Only Analytical Reagent (AR) grade chemicals and reagents to be used.

**9.2** Ultrapure water to be used for preparation of stock standards and reagents.

**9.3** Glassware to be used for sample processing and storage are to be soaked in 10 percent nitric acid followed by rinsing with ultrapure water before use, to prevent sorption of elements on the walls of the container.

**9.4** Analyse duplicate samples, reagent/process blanks, analytical blanks and known standards during the process for quality control.

### 10 PRECAUTIONS

#### 10.1 Effect of pH on Fluorimeter Response

The pH of the processed sample should be slightly acidic or neutral. The pH of the solution should not be too acidic. Also, highly acidic sample will hinder the complexation of uranium with the SPP. Alkaline

pH must be avoided because iron and calcium will be precipitated and hence the chance of wall absorption leading to lower transparency of fluorescence photons. After addition of the buffer, the pH of the solution should be in the range of 6 to 7 for maximum fluorescence response.

#### 10.2 Effect of TDS on Fluorimeter Response

Fig. 3 shows the effect of TDS on fluorimeter response. Most often high levels of TDS are caused by the presence of K, Cl, Na, etc. Beyond 1 500 ppm of TDS the fluorimetric response of the uranium analyser decreases due to decrease in transparency of sample to fluorimetric photons. Therefore, in case of high TDS content in the sample, precautions should be taken during measurement of uranium in water samples directly by fluorimetric techniques. The sample needs to be diluted accordingly and analysed to get proper fluorimetric response with respect to uranium concentration.




FIG. 3 EFFECT OF TDS ON LED FLUORIMETER RESPONSE

### 10.3 Effect of Temperature and Dust Free Environment

The method involves fluorescence emission and detection of emitted photons via photocathode of a photo-multiplier tube, both temperatures dependent. Hence the surrounding temperature should remain constant for a steady result. The laboratory should be dust free and the tissue papers used for swiping cuvette should be fiber free to avoid baffling scattering in the cuvette.

### 11 BIBLIOGRAPHY:

a) Mohapatra, S., Sahoo, S. K., Dubey, J. S.,

Patra, A. C., Thakur, V. K., Tripathy, S. K., Sagar, D. V., Godbole, S. V., Ravi, P. M., Tripathi, R. M., 2015, Characterization of uranium and its progenies in drinking water and assessment of dose to public around a NHBRA, Odisha, India. *J Radioanal Nucl Chem*, 303, 601–613;

b) Chakrabarty, A., Mohapatra, S., Tripathi, R. M., Puranik, V. D., and Kushwaha, H. S., 2010, Quality control of uranium concentration measurements. *Accreditation and Quality Assurance*, 15(2), 119–123.

**ANNEX A**

(Clause 7.1)

**URANIUM ANALYSIS IN SEAWATER SAMPLES BY LED FLUORIMETRY**

**A-0 GENERAL**

In sea water samples the interference due to high concentration of chloride ion decreases the uranyl fluorescence in the aqueous solution of the sample. The fluorescent response from a given uranium level is reduced by nearly 50 percent at 500 ppm chloride.

**A-1 REAGENTS**

**A-1.1  $K_2S_2O_8$**

**A-1.2 10 N NaOH Solution**

**A-1.3 10 Percent  $HNO_3$  Solution**

**A-3 PROCEDURE**

**A-3.1** Pipette 5 ml sample into a 150 ml beaker.

**A-3.2** Add 1g to 2 g of  $K_2S_2O_8$ , followed by the addition of 20 ml of deionised water.

**A-3.3** The solution is boiled to dryness in a hotplate (20 min to 30 min). White residue appears, then melts with evolution of white fumes.

**A-3.4** Remove beaker from hot plate and add 40 ml of deionised water after cooling to room temperature.

**A-3.5** Check pH after shaking well for dissolution. In case of detection of acidic pH, 2 to 3 drops of NaOH have to be added to make pH slightly basic (7 to 8), to ensure neutralization of excess persulfate.

**A-3.6** Add few drops of 10 percent  $HNO_3$  to make solution slightly acidic and the make-up volume up to 50 ml.

**A-3.7** Cool the solution to room temperature.

**A-3.8** Analyse by standard addition method.

**A-4 PRECAUTION**

No solids should be left when the neutralization is completed. Since the residue is primarily due to potassium bisulphate, failure to follow this precaution will result in low final pH and hence low response during analysis.



## ANNEX B

(Foreword)

## COMMITTEE COMPOSITION

Nuclear Materials for Peaceful Applications Sectional Committee, CHD 30

<i>Organization</i>	<i>Representative(s)</i>
Bhabha Atomic Research Centre, Mumbai	DR D. K. ASWAL ( <b>Chairperson</b> )
Amity Institute of Nuclear Science & Technology, Noida	DR ALPANA GOEL PROF H. S. SHARMA ( <i>Alternate</i> )
Atomic Energy Regulatory Board, Mumbai	SHRI S. K. PAWAR
Atomic Mineral Directorate, Hyderabad	DR T. S. SUNILKUMAR DR B. MAHALAKSHMI ( <i>Alternate</i> )
Bhabha Atomic Research Centre, Mumbai	DR B. K. SAPRA DR S. K. JHA ( <i>Alternate</i> )
Bharat Heavy Electricals Limited, Trichy	SHRI R. RAVI VARMA SHRI M. ARUN KUMAR ( <i>Alternate</i> )
Board of Radiation and Isotope Technology (BRIT), New Delhi	SHRI N. JAYACHANDRAN DR VIJAY KADWAD ( <i>Alternate</i> )
Defence Research & Development Organisation, Jodhpur	DR PRADEEP NARAYAN
Electronics Corporation of India Ltd, Hyderabad	SHRI P. C. SWAIN SHRI P. SUDHAKAR ( <i>Alternate</i> )
Heavy Water Board, Mumbai	MS ANANYA VERMA
Indian Institute of Management, Ahmedabad	PROF M. P. RAM MOHAN
Indian Institute of Technology, Kanpur	PROF SACHCHIDANAND TRIPATHI
Indian Rare Earths Ltd Research Centre, Mumbai	SHRI D. SINGH DR B. R. MISHRA ( <i>Alternate</i> )
Indira Gandhi Centre for Atomic Research, Kalpakkam	DR B. VENKATARAMAN DR K. SANKARAN ( <i>Alternate</i> )
Institute of Nuclear Medicine and Allied Sciences, New Delhi	DR ARUNA KAUSHIK SHRI PRADEEP GOSWAMI ( <i>Alternate I</i> ) SHRI SUKHVIR SINGH ( <i>Alternate II</i> )
Larsen and Tubro Ltd, Mumbai	SHRI VIKRAM SEHGAL SHRI ANIL PARAB ( <i>Alternate</i> )
Ministry of Defence (DGQA), Pune	LT COL PANKAJ AHUJA LT COL AMAN SODHI ( <i>Alternate</i> )
Nuclear Fuel Complex, Hyderabad	DR B. N. MURTHY SHRI Y. BALAJI RAO ( <i>Alternate</i> )
Nuclear Power Corporation of India Ltd, Mumbai	SHRI K. VENKATARAMANA MS VIBHA HARI ( <i>Alternate</i> )
Tata Memorial Centre, Mumbai	DR R. KINHIKAR DR A. JHA ( <i>Alternate</i> )
Uranium Corporation of India Ltd, Singhbhum	DR C. K. ASNANI DR P. K. TAMRAKAR ( <i>Alternate</i> )

<i>Organization</i>	<i>Representative(s)</i>
In Personal Capacity (Flat no. B117, Block 18, Kendriya Vihar CHS, Sector 38, Nerul (West), Navi Mumbai-400706)	SHRI BHASKAR PANDIT
BIS Directorate General	A. K. LAL SCIENTIST 'F'/SENIOR DIRECTOR AND HEAD (CHEMICAL) [REPRESENTING DIRECTOR GENERAL, ( <i>Ex-officio</i> )]

*Member Secretary*

SHRI PUSHPENDRA KUMAR  
SCIENTIST 'B'/ASSISTANT DIRECTOR  
(CHEMICAL), BIS

Subcommittee Responsible for Revision of IS 14194 Part 3 CHD 30 : SC 2

<i>Organization</i>	<i>Representative(s)</i>
Bhabha Atomic Research Centre, Mumbai	DR S. K. JHA ( <b><i>Convenor</i></b> )
Atomic Energy Regulatory Board, Mumbai	SHRI S. K. PAWAR DR R. M. NEHRU ( <i>Alternate</i> )
Bhabha Atomic Research Centre, Mumbai	DR ADITI C. PATRA SHRI GOPAL P. VERMA ( <i>Alternate I</i> ) SHRI S. K. SAHU ( <i>Alternate II</i> )
Board of Radiation and Isotope Technology (BRIT), Mumbai	DR AMIT SHRIVASTAVA SHRI JACOB SEBASTIAN ( <i>Alternate</i> )
Defense Research Development Organization, Delhi	DR PRADEEP NARAYAN
Indian Rare Earth Limited (Health Physics Unit), Mumbai	SHRI ABINASH SAHU
Uranium Corporation of India Ltd (Health Physics Unit), Jharkhand	DR V. N. JHA



## Bureau of Indian Standards

BIS is a statutory institution established under the *Bureau of Indian Standards Act, 2016* to promote harmonious development of the activities of standardization, marking and quality certification of goods and attending to connected matters in the country.

### Copyright

BIS has the copyright of all its publications. No part of these publications may be reproduced in any form without the prior permission in writing of BIS. This does not preclude the free use, in the course of implementing the standard, of necessary details, such as symbols and sizes, type or grade designations. Enquiries relating to copyright be addressed to the Head (Publication & Sales), BIS.

### Review of Indian Standards

Amendments are issued to standards as the need arises on the basis of comments. Standards are also reviewed periodically; a standard along with amendments is reaffirmed when such review indicates that no changes are needed; if the review indicates that changes are needed, it is taken up for revision. Users of Indian Standards should ascertain that they are in possession of the latest amendments or edition by referring to the website- [www.bis.gov.in](http://www.bis.gov.in) or [www.standardsbis.in](http://www.standardsbis.in).

This Indian Standard has been developed from Doc No.: CHD 30 (21930).

### Amendments Issued Since Publication

Amend No.	Date of Issue	Text Affected

## BUREAU OF INDIAN STANDARDS

### Headquarters:

Manak Bhavan, 9 Bahadur Shah Zafar Marg, New Delhi 110002  
Telephones: 2323 0131, 2323 3375, 2323 9402

Website: [www.bis.gov.in](http://www.bis.gov.in)

### Regional Offices:

	Telephones
Central : 601/A, Konnectus Tower -1, 6 <sup>th</sup> Floor, DMRC Building, Bhavbhuti Marg, New Delhi 110002	{ 2323 7617
Eastern : 8 <sup>th</sup> Floor, Plot No 7/7 & 7/8, CP Block, Sector V, Salt Lake, Kolkata, West Bengal 700091	{ 2367 0012 2320 9474
Northern : Plot No. 4-A, Sector 27-B, Madhya Marg, Chandigarh 160019	{ 265 9930
Southern : C.I.T. Campus, IV Cross Road, Taramani, Chennai 600113	{ 2254 1442 2254 1216
Western : Plot No. E-9, Road No.-8, MIDC, Andheri (East), Mumbai 400093	{ 2821 8093

**Branches :** AHMEDABAD. BENGALURU. BHOPAL. BHUBANESHWAR. CHANDIGARH. CHENNAI. COIMBATORE. DEHRADUN. DELHI. FARIDABAD. GHAZIABAD. GUWAHATI. HIMACHAL PRADESH. HUBLI. HYDERABAD. JAIPUR. JAMMU & KASHMIR. JAMSHEDPUR. KOCHI. KOLKATA. LUCKNOW. MADURAI. MUMBAI. NAGPUR. NOIDA. PANIPAT. PATNA. PUNE. RAIPUR. RAJKOT. SURAT. VISAKHAPATNAM.