
पेट्रोलियम एवं इसके उत्पादों की परीक्षण पद्धतियाँ

[P : 169]

क्लोरीन और ब्रोमीन सामग्री का निर्धारण —
वेवलेंथ फैलानेवाला एक्स-रे प्रतिदीप्ति स्पेक्ट्रोमेट्री

Methods of Test for Petroleum and Its Products

[P : 169]

Determination of Chlorine and Bromine Content —
Wavelength-Dispersive X-Ray Fluorescence Spectrometry

ICS 75.080

© BIS 2018



भारतीय मानक ब्यूरो

BUREAU OF INDIAN STANDARDS

मानक भवन, 9 बहादुरशाह ज़फर मार्ग, नई दिल्ली-110002

MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG

NEW DELHI-110002

www.bis.gov.in www.standardsbis.in

NATIONAL FOREWORD

The Indian Standard [P : 169] which is identical with ISO 15597 : 2001 'Petroleum and related products — Determination of chlorine and bromine content — Wavelength-dispersive x-ray fluorescence spectrometry' issued by the International Organization for Standardization (ISO) was adopted by the Bureau of Indian Standards on the recommendation of the Methods of Sampling and Test for Petroleum, Its Products, Gaseous Fuels and Lubricants Sectional Committee and approval of the Petroleum, Coal and Related Products Division Council.

The text of ISO Standard has been approved as suitable for publication as an Indian Standard without deviations. Certain conventions are however not identical to those used in Indian Standards. Attention is particularly drawn to the following:

- a) Wherever the words 'International Standard' appear referring to this standard, they should be read as 'Indian Standard'.
- b) Comma (,) has been used as a decimal marker while in Indian Standards, the current practice is to use a point (.) as the decimal marker.

The technical committee has reviewed the provisions of the following International Standards referred in this adopted standard and has decided that they are acceptable for use in conjunction with this standard:

<i>International Standard</i>	<i>Title</i>
ISO 3170 : 1988	Petroleum liquids — Manual sampling
ISO 3171 : 1988	Petroleum liquids — Automatic pipeline sampling
ISO 4259 : 1992	Petroleum products — Determination and application of precision data in relation to methods of test

For tropical countries like India, the standard temperature and the relative humidity shall be taken as $27 \pm 2^{\circ}\text{C}$ and 65 ± 5 percent, respectively.

In reporting the result of a test or analysis made in accordance with this standard, if the final value, observed or calculated, is to be rounded off, it shall be done in accordance with IS 2 : 1960 'Rules for rounding off numerical values (*revised*)'.

Indian Standard

**METHODS OF TEST FOR
PETROLEUM AND ITS PRODUCTS**

[P : 169]

**DETERMINATION OF CHLORINE AND BROMINE CONTENT —
WAVELENGTH-DISPERSIVE X-RAY FLUORESCENCE SPECTROMETRY**

WARNING — The use of this International Standard may involve hazardous materials, operations and equipment. This International Standard does not purport to address all of the safety problems associated with its use. It is the responsibility of the user of this International Standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

1 Scope

This International Standard specifies a method for the determination of the chlorine and bromine content of liquid petroleum products, synthetic oils and fluids, and additives for petroleum products (including used oils) that are soluble in organic solvents of negligible or accurately known chlorine/bromine content. The method is applicable to products or additives having chlorine contents in the range 0,0005 % (*m/m*) to 0,1000 % (*m/m*), and bromine contents in the range 0,0010 % (*m/m*) to 0,1000 % (*m/m*). Other elements do not generally interfere, although lead may interfere at contents above 0,1500 % (*m/m*) (see note 2).

NOTE 1 For the purposes of this International Standard, the term “% (*m/m*)” is used to represent the mass fraction of a material.

NOTE 2 Used lubricants may pose particular problems due to the range of potentially interfering elements at relatively high concentrations. For used lubricants generally, the lower limit of sensitivity may be 0,0050 % (*m/m*) even when the provisions of the last paragraph of 9.3 are applied.

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this International Standard. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 3170:1988, *Petroleum liquids — Manual sampling*.

ISO 3171:1988, *Petroleum liquids — Automatic pipeline sampling*.

ISO 4259:1992, *Petroleum products — Determination and application of precision data in relation to methods of test*.

3 Principle

The test portion and a bismuth solution as internal standard are mixed in a given mass ratio and exposed, in a sample cell, to the primary radiation of an X-ray tube.

The count rates of the chlorine $K\alpha$ at 0,4729 nm and bismuth $M\beta$ at 0,4909 nm, or bromine $K\alpha$ at 0,104 1 nm and bismuth $L\alpha$ at 0,114 4 nm fluorescence thus excited, and the count rate of the background radiation at 0,480 7 nm or 0,108 5 nm, are measured, and the ratio of these net count rates is calculated. The chlorine and/or bromine content of the sample is determined from calibration curves prepared on the basis of chlorine and/or bromine calibration standards.

4 Reagents and materials

4.1 **White oil (light paraffin oil)**, high purity grade, sulfur content 1 mg/kg maximum.

4.2 **Chlorine compound**, 1-chlorooctane or another oil-soluble chlorine compound of accurately known chlorine content, used for the preparation of the primary standards. The chlorine content shall be accurately known to the nearest 0,01 % (*m/m*).

4.3 **Bromine compound**, 1,1,2,2-tetrabromoethane or another oil-soluble bromine compound of accurately known bromine content, used for the preparation of the primary standards. The bromine content shall be accurately known to the nearest 0,01 % (*m/m*).

CAUTION — 1,1,2,2-tetrabromoethane is extremely toxic by inhalation and ingestion. Appropriate precautions for the handling operations of opening the container and weighing shall be followed.

4.4 **Certified reference materials**, obtained from a national standards body or accredited supplier with a range of certified chlorine and/or bromine contents for the production of calibration curves for routine analysis.

4.5 **Bismuth compound**, triphenylbismuth, of minimum purity 98 %.

4.6 **2-ethylhexanoic acid**, of minimum purity 98 %.

5 Apparatus

5.1 **Wavelength-dispersive X-ray fluorescence spectrometer**, any suitable spectrometer that allows the count rates of the Cl- $K\alpha$, Br- $K\alpha$, Bi- $M\beta$ and Bi- $L\alpha$ X-ray fluorescence to be measured, provided that the design incorporates the general features given in Table 1. It shall be set up according to the manufacturer's instructions.

Table 1 — General requirements of the spectrometer

Component	Requirement
Anode	Rhodium, scandium, chromium, or any other tube anode that allows the counting times to be adjusted to achieve the required precision
Collimator (if used)	Coarse for chlorine, narrow for bromine
Analysing crystal	Germanium for chlorine, lithium fluoride (LiF) for bromine, or any other crystal suitable for the required dispersion of the wavelengths in Table 2 within the angular range of the spectrometer
Optical path	Helium
Cell window	Polyester or polypropylene film, chlorine- and bromine-free, thickness 2 μm to 6 μm
Detector	Proportional counter with pulse-height analyser. For bromine, a scintillation counter with pulse-height analyser is preferred

5.2 **Analytical balance**, capable of weighing accurately to the nearest 0,1 mg.

5.3 **Homogenizer**, of the non-aerating, high-speed shear type, or **heatable magnetic** or **ultrasonic stirrer**.

5.4 **Filters**, of sintered glass, with a pore size of 10 μm to 60 μm .

5.5 Flasks, of 25 ml to 100 ml capacity, narrow-necked, conical, made of borosilicate glass, and fitted with ground-glass stoppers.

6 Samples and sampling

6.1 Unless otherwise specified, samples shall be taken in accordance with the procedures specified in ISO 3170 or ISO 3171.

6.2 Test portions from the samples shall be drawn after thorough mixing and subdivision. Heat viscous samples to a temperature which renders the sample fluid, and homogenize, using the homogenizer or stirrer (5.3) as necessary.

NOTE For the purpose of this procedure, the term "sample" includes solutions prepared from additives, semi-solid or solid petroleum products that have been appropriately pre-treated and/or diluted.

7 Calibration solutions

7.1 General

Use either certified reference materials (4.4) or primary standards prepared from chlorine compounds (4.2) and/or bromine compounds (4.3) dissolved in white oil (4.1) as a basis for the preparation of the appropriate range of stock solutions.

7.2 Preparation of stock solutions

7.2.1 Chlorine and bromine stock solutions

Weigh, to the nearest 0,1 mg, a quantity of chlorine compound (4.2) or bromine compound (4.3) to prepare stock solutions of approximately 0,10 % by mass of chlorine or bromine, calculated to the nearest 0,001 % (*m/m*), and dissolve in white oil (4.1) at room temperature. Mix the contents thoroughly using a homogenizer (5.3), and store in a glass-stoppered flask (5.5).

It is recommended that a polytetrafluorethylene (PTFE) or glass-coated magnetic stirrer and stirring device be used to mix the contents of the flask.

Calculate the exact chlorine and/or bromine content, w_2 , in percent by mass, to three decimal places, in each case from the amounts of white oil and compound used as follows:

$$w_2 = \frac{m_c \times w_1}{m_c + m_0} \quad (1)$$

where

m_c is the mass of the chlorine or bromine compound, expressed in grams;

w_1 is the chlorine or bromine content of the compound, expressed in percent by mass;

m_0 is the mass of white oil, expressed in grams.

7.2.2 Bismuth solution

Weigh, to the nearest 0,1 mg, a quantity of triphenylbismuth (4.5) sufficient to make a 1 % by mass solution in white oil. Dissolve the triphenylbismuth in white oil by stirring and warming to a temperature not exceeding 80 °C. Add 5 % (*m/m*) 2-ethylhexanoic acid (4.6) to the solution. Remove any residual turbidity by filtering the gently heated mixture through the filter (5.4). Store in a tightly stoppered flask or bottle.

When protected against moisture, the bismuth solution can be kept for several months. Any turbidity should be taken as an indication that the solution is no longer suitable for use.

7.3 Preparation of standard solutions

Prepare standard solutions of chlorine and/or bromine contents of 0,000 5 % (m/m), 0,001 0 % (m/m), 0,002 0 % (m/m), 0,005 0 % (m/m), 0,010 0 % (m/m), 0,025 0 % (m/m) and 0,050 0 % (m/m) from the stock solutions (7.2.1), by weighing, to the nearest 0,1 mg, a quantity of the stock solution to produce approximately 25 g of standard solution, into a flask (5.5), and diluting this with white oil (4.1). Reweigh to the nearest 0,1 mg and mix thoroughly at room temperature. Calculate the content of chlorine or bromine, w_3 , in percent by mass, in each solution to the nearest 0,000 01 % (m/m) by means of equation (2) below. Transfer the solutions to tightly stoppered dark glass bottles with the content recorded on each bottle.

$$w_3 = \frac{w_2 (m_2 - m_1)}{m_3 - m_1} \quad (2)$$

where

m_1 is the mass of the flask, expressed in grams;

m_2 is the mass of the flask plus stock solution, expressed in grams;

m_3 is the mass of the flask plus stock solution plus white oil, expressed in grams.

7.4 Preparation of calibration solutions

Weigh a series of flasks (5.5) and add a quantity of each standard solution (7.3) to separate flasks. Also add a quantity of each stock solution (7.2.1) and of the white oil (4.1) to additional separate weighed flasks. Reweigh. Mix each of these with bismuth solution (7.2.2) in the ratio 10:1. Mix thoroughly at room temperature. The quantity of each resulting solution shall be sufficient to adequately fill the sample cells for the spectrometer.

NOTE Normally, 25 g \pm 0,1 g of solution plus 2,5 g \pm 0,01 g of bismuth solution gives an adequate volume.

7.5 Storage of standards

Store certified reference materials in accordance with the instructions of the certifying organization, and use within the time-scale specified.

Store standards prepared from white oil and chlorine and/or bromine compounds in dark glass-stoppered bottles in a cool dark place.

8 Calibration

8.1 General

After the spectrometer (5.1) has been set up and checked, purge the optical path thoroughly with helium.

8.2 Measurements

Transfer each of the calibration solutions (7.4) to a sample cell and, in a sequence of increasing chlorine or bromine content, place them in the spectrometer for exposure to the primary radiation.

Measure the count rates, I , at the wavelengths specified in Table 2, and just one background radiation at an appropriate wavelength for the X-ray tube being used.

NOTE Lead interferes with background measurements on the short wavelength side of the bromine line.

Keep the total duration of the radiation as short as possible, measuring the X-ray fluorescence radiation for each solution for chlorine and/or bromine followed by bismuth. For the chlorine or bromine radiation, do not exceed a measurement time of 100 s, or 40 s if an end-window tube is used.

Choose a measurement procedure that results in at least 50 000 pulses for the Bi-M β or Bi-L α radiation during the total measurement period.

Table 2 — Wavelengths of the X-ray fluorescence and background radiation

Radiation		Wavelength nm
Chlorine	Cl-K α	0,472 9
Bismuth (Cl)	Bi-M β	0,490 9
Background, Cl	B, Cl	0,480 7
Bromine	Br-K α	0,104 1
Bismuth (Br)	Bi-L α	0,114 4
Background, Br	B, Br	0,108 5

Calculate the net count ratios, $R_{0,Cl}$ and $R_{0,Br}$ from equation (3) or equation (4).

$$R_{0,Cl} = \frac{I_{Cl} - I_{B,Cl}}{I_{Bi,Cl} - I_{B,Cl}} \quad (3)$$

$$R_{0,Br} = \frac{I_{Br} - I_{B,Br}}{I_{Bi,Br} - I_{B,Br}} \quad (4)$$

where

I_{Cl} is the count rate for Cl-K α ;

$I_{Bi,Cl}$ is the count rate for Bi-M β ;

$I_{B,Cl}$ is the count rate for background radiation (chlorine);

I_{Br} is the count rate for Br-K α ;

$I_{Bi,Br}$ is the count rate for Bi-L α ;

$I_{B,Br}$ is the count rate for background radiation (bromine).

8.3 Calibration curves

Plot the count ratio, R_0 , of the individual calibration solutions as a function of the chlorine and/or bromine content, to give calibration curves for chlorine and/or bromine.

NOTE A polynomial equation may be required for mathematical analysis of the experimental data.

8.4 Checking

Check the calibration curves regularly, and after each fresh preparation of bismuth solution. In regular use, check at least two points on each calibration curve not less frequently than every six months. If the check result differs from the curve by more than the repeatability of this International Standard, construct a new calibration curve.

9 Procedure

9.1 Ensure that all laboratory glassware is free from chlorine or bromine, and that the laboratory air does not contain any vapour from halogenated solvents.

9.2 Weigh, to the nearest 0,1 mg, $25 \text{ g} \pm 0,1 \text{ g}$ of the sample to be analysed into a flask (5.5), and add $2,5 \text{ g} \pm 0,01 \text{ g}$ of the bismuth solution (7.2.2), weighed to the same accuracy. Stopper immediately and homogenize thoroughly using the homogenizer (5.3) and gently heat to a temperature not exceeding 50 °C.

9.3 Cool the solution to room temperature, transfer it to a sample cell and place it in the spectrometer for exposure to the primary radiation. Measure the count rates, I , at the specified wavelengths, together with the background radiations. Measure the count rates, I_{Bi} , for the appropriate bismuth lines and calculate the count ratio, $R_{0,\text{Cl}}$ or $R_{0,\text{Br}}$, from equation (3) and/or equation (4).

If bubble formation is found to be a problem, and if both chlorine and bromine contents are required, divide the solution into two sample cells to reduce the overall radiation period, and determine the concentrations separately.

If interfering elements, e.g. lead at contents above 0,15 % (m/m), are present, then the use of a primary collimator with higher resolution than would normally be used for the measurement of Cl-K α is recommended.

10 Calculation

Read the chlorine and/or bromine content from the calibration curves constructed in accordance with 8.3. If the chlorine or bromine content is above that specified in the scope of this International Standard, dilute the sample with white oil (4.1) and repeat the procedure specified in 9.2 and 9.3.

NOTE Many modern instruments contain a microprocessor that retains the calibration curve and produces a digital readout.

11 Expression of results

Report the chlorine and/or bromine content of the sample to the nearest 0,000 1 % (m/m), or to the nearest 1 mg/kg.

12 Precision

12.1 General

The precision, as determined by statistical examination, in accordance with ISO 4259, of interlaboratory test results is given in 12.2 and 12.3.

12.2 Repeatability

The difference between two test results obtained by the same operator with the same apparatus under constant operating conditions on identical test material would in the long run, in the normal and correct operation of the test method, exceed the values given in Table 3 in only one case in twenty.

12.3 Reproducibility

The difference between two single and independent test results obtained by different operators working in different laboratories on identical test material would in the long run, in the normal and correct operation of the test method, exceed the values given in Table 3 in only one case in twenty.

Table 3 — Precision values

Halogen content % (m/m) [mg/kg]	Repeatability % (m/m) [mg/kg]	Reproducibility % (m/m) [mg/kg]
Chlorine		
0,000 5 to 0,001 0 [5 to 10]	0,000 15 [1,5]	0,000 3 [3]
> 0,001 0 to 0,005 0 [> 10 to 50]	0,000 3 [3]	0,000 6 [6]
> 0,005 0 to 0,010 0 [> 50 to 100]	0,000 5 [5]	0,001 0 [10]
> 0,010 0 to 0,100 0 [> 100 to 1 000]	3 % relative ^a	10 % relative ^a
Bromine		
0,001 0 to 0,005 0 [10 to 50]	0,000 2 [2]	0,000 5 [5]
> 0,005 0 to 0,010 0 [> 50 to 100]	0,000 3 [3]	0,001 0 [10]
> 0,010 0 to 0,100 0 [> 100 to 1 000]	3 % relative ^a	10 % relative ^a
^a Based on the mean value calculated from two results.		

13 Test report

The test report shall contain at least the following information:

- a) a reference to this International Standard;
- b) the type and complete identification of the product tested;
- c) the result of the test (see clause 11);
- d) any deviation, by agreement or otherwise, from the procedures specified;
- e) the date of the test.

Bureau of Indian Standards

BIS is a statutory institution established under the *Bureau of Indian Standards Act, 2016* to promote harmonious development of the activities of standardization, marking and quality certification of goods and attending to connected matters in the country.

Copyright

BIS has the copyright of all its publications. No part of these publications may be reproduced in any form without the prior permission in writing of BIS. This does not preclude the free use, in course of implementing the standard, of necessary details, such as symbols and sizes, type or grade designations. Enquiries relating to copyright be addressed to the Director (Publications), BIS.

Review of Indian Standards

Amendments are issued to standards as the need arises on the basis of comments. Standards are also reviewed periodically; a standard along with amendments is reaffirmed when such review indicates that no changes are needed; if the review indicates that changes are needed, it is taken up for revision. Users of Indian Standards should ascertain that they are in possession of the latest amendments or edition by referring to the latest issue of 'BIS Catalogue' and 'Standards: Monthly Additions'.

This Indian Standard has been developed from Doc No.: PCD 01 (10587).

Amendments Issued Since Publication

Amendment No.	Date of Issue	Text Affected

BUREAU OF INDIAN STANDARDS

Headquarters:

Manak Bhavan, 9 Bahadur Shah Zafar Marg, New Delhi 110002

Telephones: 2323 0131, 2323 3375, 2323 9402

Website: www.bis.gov.in

Regional Offices:

Telephones

Central : Manak Bhavan, 9 Bahadur Shah Zafar Marg
NEW DELHI 110002

{ 2323 7617
2323 3841

Eastern : 1/14, C.I.T. Scheme VII M, V.I.P. Road, Kankurgachi
KOLKATA 700054

{ 2337 8499, 2337 8561
2337 8626, 2337 9120

Northern : Plot No. 4-A, Sector 27-B, Madhya Marg, CHANDIGARH 160019

{ 26 50206
265 0290

Southern : C.I.T. Campus, IV Cross Road, CHENNAI 600113

{ 2254 1216, 2254 1442
2254 2519, 2254 2315

Western : Manakalaya, E9 MIDC, Marol, Andheri (East)
MUMBAI 400093

{ 2832 9295, 2832 7858
2832 7891, 2832 7892

Branches: AHMEDABAD. BENGALURU. BHOPAL. BHUBANESWAR. COIMBATORE.
DEHRADUN. DURGAPUR. FARIDABAD. GHAZIABAD. GUWAHATI.
HYDERABAD. JAIPUR. JAMMU. JAMSHEDPUR. KOCHI. LUCKNOW. NAGPUR.
PARWANOO. PATNA. PUNE. RAIPUR. RAJKOT. VISAKHAPATNAM.

Published by BIS, New Delhi