



भारतीय मानक ब्यूरो

(उपभोक्ता मामले, खाद्य एवं सार्वजनिक वितरण मंत्रालय, भारत सरकार)

BUREAU OF INDIAN STANDARDS

(Ministry of Consumer Affairs, Food & Public Distribution, Govt. of India)

मानक भवन, 9 बहादुरशाह जफर मार्ग नई, दिल्ली-110002

Manak Bhavan, 9 Bahadur Shah Zafar Marg, New Delhi-110002

Phones: 23230131 / 23233375 / 23239402

Website: www.bis.org.in, www.bis.gov.in

DRAFT INDIAN STANDARD IN WIDE CIRCULATION

Reference : 22615

Date : 18 June 2023

TECHNICAL COMMITTEE : Anaesthetic, Resuscitation And Allied Equipment, MHD 11

To,

All concerned

Dear Madam/Sir,

The following document has been prepared by the Anaesthetic, Resuscitation And Allied Equipment Sectional Committee, MHD 11. Please [click here](#) to view the document.

Document Number : MHD 11 (22615) WC

Title of the document : Anaesthetic and Respiratory Equipment- Oropharyngeal Airways(First Revision)

Document Type : Revision of Indian Standard (IS/ISO 5364 : 2008)

This document has following salient features which may require specific attention for your valuable comments:

- 1) This Indian Standard specifies requirements for oropharyngeal airways of plastics materials and/or rubber, including those with a reinforcement insert made of plastics materials and/or metal. This Indian Standard is not applicable to metal oropharyngeal airways, nor to requirements concerning flammability of oropharyngeal airways. Flammability of oropharyngeal airways, for example, if flammable anaesthetics, electrosurgical units, or lasers are used, is a well-recognized hazard. It is addressed by appropriate clinical management, which is outside the scope of this International Standard. This Indian Standard is not applicable to supralaryngeal airways without an internal, integral sealing mechanism.*
- 2) This Indian Standard specifies requirements for oropharyngeal airways of plastics materials and/or rubber, including those with a reinforcement insert made of plastics materials and/or metal. This Indian Standard is not applicable to metal oropharyngeal airways, nor to requirements concerning flammability of oropharyngeal airways. Flammability of oropharyngeal airways, for example, if flammable anaesthetics, electrosurgical units, or lasers are*

used, is a well-recognized hazard. It is addressed by appropriate clinical management, which is outside the scope of this Indian Standard. This Indian Standard is not applicable to supralaryngeal airways without an internal, integral sealing mechanism.

Please examine the document and share your comments regarding further improvement in the document.

Last date for sharing the comments is : 18 July 2023

The comments should be shared in the prescribed template through this portal only; and the comments so received shall be taken up by the Sectional Committee for necessary action. For any other query, please write an email at mhd@bis.gov.in to the undersigned at Bureau of Indian Standard, Manak Bhawan, 9, Bahadur Shah Zafar Marg, New Delhi.

In case no comments are received, we would presume your approval of the documents. However, in case we receive any comments on the document, the same shall be put up to the Sectional Committee for necessary action.

Thanking You,

Yours faithfully,
(UNNIKRISHNAN A R)
Head (Medical Equipment and Hospital Planning Department)
Email: mhd@bis.gov.in



व्यापक परिचालन में मसौदा(दे)

हमारा सन्दर्भ : 22615

दिनांक : 18-06-2023

तकनीकी समिति : Anaesthetic, Resuscitation And Allied Equipment Sectional Committee, MHD 11

प्राप्तकर्ता : रूचि रखने वाले सभी निकाय

महोदय/या,

निम्नलिखित मसौदा तैयार किया गया है :

प्रलेख संख्या : MHD 11 (22615) WC

शीर्षक :

कृपया इस/इन मानक(को)/संशोधन(नो) के मसौदे(दो) का अवलोकन करें और अपनी सम्मतियाँ यह बताते हुए भेजें कि यदि ये मानक(को) के संशोधन(नो) के रूप में प्रकाशित हो तो इन पर अमल करने में आपके व्यवसाय अथवा कारोबार में क्या कठिनाइयां आ सकती हैं।

सम्मतियाँ भेजने की अंतिम तिथि : 18 July 2023

सम्मतियाँ, यदि कोई हों तो, कृपया यहाँ क्लिक करके ऑनलाइन पोर्टल के माध्यम से ऊपर दी गयी अंतिम तिथि तक दर्ज कराएं।

यह/ये प्रलेख भारतीय मानक ब्यूरो की वेबसाइट www.bis.gov.in पर भी उपलब्ध है/हैं।

धन्यवाद।

भवदीय/भवदिया,
विभाग प्रमुख का नाम : UNNIKRISHNAN A R
(Medical Equipment and Hospital Planning Department)
ई-मेल : mhd@bis.gov.in