

**BUREAU OF INDIAN STANDARDS
(Jaipur Branch Office)**

Our Ref: JPBO/A-84002243

20.05.2024

Sub: All India Application as per IS 407:1981 for Specification for brass tubes for general purposes (Third Revision)

JPBO has received an application from M/s MULTIMETALS LIMITED under Option 1 for grant of License as per **IS 407:1981** for the product **Specification for brass tubes for general purposes (Third Revision)**

As per the online information available on BIS website and Intranet:

1. There is no operative license for this product.
2. Minimum Marking fees is not available for this product.
3. Product manual and SIT is not available.

As no operative license is available. This may be treated as All India First Case for the above product.

CMD 2 and all BOs/ROs are therefore requested to provide any information with respect to above IS/Product which may be useful to process the application.

LPPD is requested to inform the availability of testing facility in BIS lab or any other BIS recognized laboratory and if so, testing charges may be informed.

ITSD is requested to kindly host the above information on the intranet.

Udit
20.05.24

Udit Agrawal
Scientist-B

Head JPBO *dm* 20.05.2024

Head (CMD-2)/Head (LPPD)/Head ROs/Head BOs

Copy to:

DDGC

DDG (Certification)

ITSD – For circulation on Intranet

भारतीय मानक ब्यूरो
जयपुर शाखा कार्यालय

संदर्भ: JPBO/A-84002243

20.05.2024

विषय: IS 407:1981 के अनुसार अखिल भारतीय प्रथम आवेदन

जयपुर शाखा कार्यालय को IS 407:1981 अनुसार मानक चिन्ह का उपयोग करने के लिए आवेदन प्राप्त हुआ है।


बीआईएस वेबसाइट/ इंटरनेट पर उपलब्ध सूचना से, निम्न बिंदु देखे गये है :

- 1 वर्तमान तिथि तक, उक्त उत्पाद के लिए कोई बीआईएस लाइसेंस उपलब्ध नहीं है
- 2 उक्त उत्पाद के लिए अंकन शुल्क दर उपलब्ध नहीं है


इसलिए, इस आवेदन को उत्पाद के लिये अखिल भारतीय प्रथम आवेदन माना गया है

सीएमएडी -2, सभी क्षेत्रीय कार्यालय एवं शाखा कार्यालयों से अनुरोध है कि उक्त आई एस से संबंधित कोई जानकारी उपलब्ध हो या कोई अन्य सूचना जो इस आवेदन को आगे बढ़ाने में उपयोगी हो, उपलब्ध कराने की कृपा करें

एलपीपीडी से अनुरोध है कि बी आई एस स्वीकृत प्रयोगशालाओं या अन्य प्रयोगशालों की जानकारी प्रदान करें, जहां इस उत्पाद से संबंधित टेस्ट किए जा सकते हैं।


20.05.24

उदित अग्रवाल
सहायक निदेशक

प्रमुख (जशाका) 
20.05.2024

प्रमुख (CMD 2)/प्रमुख (एलपीपीडी)/ प्रमुख (ROs)/ प्रमुख (BOs)

प्रतिलिपि:

डीजीजीसी

डीडीजी (प्रमाणीकरण)

आईटीएसडी