



**BUREAU OF INDIAN STANDARDS**  
**(Lab Name)**



सन्दर्भ : उक्षेका प्रयोगशाला/ केमि/IS 10810 (P-4)

Date: 13-09-2021

Ref: NROL/Chem/IS 10810(P-4)

उत्तर क्षेत्रीय कार्यालय प्रयोगशाला मोहाली में पर-सल्फेट परीक्षण की आंशिक सुविधा का निर्माण:-

**Creation of Test facilities of Per-sulphate test as per IS 10810(P-4) at NROL Mohali as detailed below:-**

आइ एस क्रमांक तथा उत्पाद / IS No. and Product	:	<b>IS 694</b> Polyvinyl chloride insulated unsheathed and sheathed cables/cords with rigid and flexible conductor <b>IS 2593</b> , Flexible Cables for Miners' Cap-lamps <b>IS 17048</b> , Halogen free flame retardant (HFFR) cables for working voltages up to and including 1 100 Volts
	:	
Test (s) with Clause No. for which facility has been created	:	पर-सल्फेट परीक्षण /Persulphate test
Type/Size/Grade/Variety for which test facility has been created	:	आइ एस 10810(P-4) As per IS 10810(P-4)
Test Section	:	रसायन प्रयोगशाला / Chemical Laboratory
Major equipment procured for creation of test facility	:	मैंडरल /Mandrels Chemical
Status of facility created	:	आंशिक/ Partial
Test (s) with Clause No. for which facility yet to be created for complete testing	:	लागू नहीं / Not applicable
Sample size	:	5 मीटर केबल वायर / 5 meter cable wire
Testing capacity/month	:	10
Remarks	:	Remaining testing carried out in Electrical lab

OIC (testing section)

Head (Lab)

Issued in consent with competent authority

ITSD with request( To upload in intranet)