

भारतीय मानक ब्यूरो
(प्रयोगशाला मान्यता और प्रबंधन विभाग)

संदर्भ: एलआरएस/ ओएसएल-6172206

22-05-2023

विषय: मैं ऐ.आई.सी.-ऐ.एम.टी.ज़ेड मेडीवल्ले इंकूबेशन काउंसिल (मेडीवल्ले लैबोरेटरी) (ओएसएल-6172206) की मान्यता को अनुदान ।

सक्षम प्राधिकारी के अनुमोदन अनुसार मैं ऐ.आई.सी.-ऐ.एम.टी.ज़ेड मेडीवल्ले इंकूबेशन काउंसिल (मेडीवल्ले लैबोरेटरी), ऐ.एम.टी.ज़ेड एड्मिनिसट्रेशन बिल्डिंग, ऐ.एम.टी.ज़ेड कैम्पस, प्रगति मेदान, वी.एम. स्टील प्रोजेक्ट, विशाखपटनम, आंध्रा प्रदेश- 530031 (ओएसएल-6172206) को भारतीय मानक ब्यूरो की प्रयोगशाला मान्यता योजना के अंतर्गत दिनांक 19/05/2023 से 3 साल की अवधि के लिए मान्यता अनुदान दिया गया है।

प्रयोगशाला मान्यता प्राप्त क्षेत्र अनुबंध - 1 में दिया गया है।

पुट्टी भानु कृष्णदेव
वैज्ञानिक- बी (एल आर एम डी)

प्रमुख (एल आर एम डी)

प्रतिलिपि - सभी क्षेत्रीय एवं शाखा कार्यालय / प्रयोगशालाएं / सी एम डी -1 / सी एम डी -2 / सी एम डी -3 के लिए प्रमुख आई टी एस : कृपया संलग्न किए गए प्रपत्र को भारतीय मानक ब्यूरो के इंटरनेट पर अपलोड करें।

Bureau of Indian Standards

(Laboratory Recognition and Management Department)

Our Ref: LRMD/LRS/OSL-(6172206)

22-05-2023

Subject: Grant of Recognition to M/s AIC-AMTZ Medivalley Incubation Council (Medivalley laboratory) (OSL Code-6172206).

As approved by the Competent Authority M/s AIC-AMTZ Medivalley Incubation Council (Medivalley laboratory), AMTZ Administrative Building, AMTZ Campus, Pragati Maidan VM Steel Project, Visakhapatnam, Andhra Pradesh-530031 (OSL Code-6172206) has been granted recognition for three year w.e.f. 19-05-2023 under the Laboratory Recognition Scheme of BIS.

The scope of Recognition of the lab is enclosed as **Annex-1**

Putti Bhanu Krishnadev
Scientist-B (LRMD)

Head, (LRMD)

CC: All ROs/BOs/BIS Labs/CMD-I/CMD-II/CMD-III

Head ITS, BIS HQ-With a request to upload this circular on BIS internet

OSL CODE 6172206

NAME AIC-AMTZ Medivalley Incubation Council
 ADDRESS (Medivalley laboratory)AMTZ Administrative Blding
 ,AMTZ Campus, Pragati Maidan,V.M Steel Project
 Visakhapatnam 530031
 Andhra Pradesh

CONTACT :

TELEPHONE: 9676126445

FAX:

E-mail surya.ch@medivalley-aic.in

REMARKS

IS Number	Product	Testing Charge	Remark	
1	IS 13450(Part 1/Sec 2): 2018	Medical electrical equipment: Part 1 general requirements for the basic safety and essential performance: Sec 2 collateral standard: electromagnetic disturbances - Requirements and tests (First Revision)	Rs. 4,00,000/-	Exclusion: Cl.4.3 (Mains terminal disturbance voltage (Conducted EMISSIONS) CISPER16, Electromagnetic radiation disturbance (radiated Emissions) CISPER 11, Harmonics Current Emissions IEC 61000-3-2 & 3-3, Voltage changes, Voltage Fluctuations and flickers EMISSIONS IEC 61000-3-3, Immunity to proximity fields from RF wireless Communications Equipment IEC 61000-4- 3, Power Frequency magnetic field IMMUNITY IEC61000-4- 8, Voltages dips ,Short Interruptions and Voltages Variations IMMUNITY IEC 61000-4- 11, Radiated RF Electromagnetic Field Immunity IEC61000-4- 3), Cl.7.1 (PROTECTION OF RADIO SERVICES AND OTHER EQUIPMENTS), Table-5 (Voltage dips IEC 61000-4-11, Voltage Interruptions IEC 61000- 4-11), Table-4 (Radiated RF Field EM Field IEC 61000-4-3, Proximity field From RF Wireless, Power Frequency Magnetic Fields) Cl.8.6 (Radiated RF Electromagnetic Field Immunity IEC61000-4- 3), Cl.7.1 (Electromagnetic radiation disturbance (radiated Emissions) CISPER 11), Cl.7.2 (Harmonics distortions.)