

				Dimension 19.8 Major dia. at small end (dle) is given 17.61 which has corrected in this P draft as 17.16. Considering pitch 1.162, value 17.16 may be correct, was there any amendment for changed size (17.61 to 17.16).		has perused the comment and informed that the value given in this draft is correct and no change is required. The Committee agreed to the proposal and decided no change in the draft.
--	--	--	--	---	--	---

The Committee approved the above document for wide circulation.

Item 5.2 Doc:MED 16 (17432) Specification for Residual pressure valves and type testing of cylinder valves incorporating residual pressure devices

The Committee noted the information as given in the agenda. Being new document, the Committee decided to recirculate the draft to the Committee members for their examination. Shri Y.K Behani agreed to provide the details relevant to be provided in the foreword of the draft before circulation to members within 10 days.

If no comments is received from the Committee members, the draft to be put into wide circulation for comments with approval of Chairman.

Item 5.3 Doc:MED 16 (16579) Draft Indian Standard on Yoke Type Valve Connections for Medical Gas Cylinders Specification (Third revision of IS 3745)

The Committee noted the information as given in the agenda. The Committee decide to put the draft into wide circulation for 60 days for comments.

Item 6 COMMENTS ON PUBLISHED INDIAN STANDARDS

6.1 IS 7302 : 2018 Valve fittings for self contained breathing apparatus SCBA and self contained underwater breathing apparatus SCUBA - Specification (First Revision)