

निर्वात् गेज के अंशशोधन की विधि
भाग 1 10^{-5} Pa से 10^{-1} Pa की दाब सीमा में सतत
प्रवाह से दाब में कटौती
(पहला पुनरीक्षण)

**Method for Calibration of
Vacuum Gauges
Part 1 Pressure Reduction by Continuous
Flow in the Pressure Range of
 10^{-5} Pa to 10^{-1} Pa
(First Revision)**

ICS 23.160

© BIS 2024



भारतीय मानक ब्यूरो
BUREAU OF INDIAN STANDARDS
मानक भवन, 9 बहादुर शाह ज़फर मार्ग, नई दिल्ली - 110002
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG
NEW DELHI - 110002

www.bis.gov.in www.standardsbis.in

FOREWORD

This Indian Standard (First Revision) was adopted by the Bureau of Indian Standards, after the draft finalized by the Chemical Engineering Plants and Related Equipment Sectional Committee had been approved by the Mechanical Engineering Divisional Council.

This standard was first published in 1976. This revision has been taken up with a view incorporating the modification found necessary as a result of experience gained in the use of this standard. Also, in this revision, the standard has been brought into the latest style and format of Indian Standards, and references to Indian Standards, wherever applicable have been updated.

International system (SI) units have been used in this standard. The relation of these units to other units is given below for guidance:

$$1 \text{ Pascal (Pa)} = 1 \text{ newton/square metre (N/m}^2\text{)}$$

$$1 \text{ torr} = 1 \text{ mm Hg} = 133.322 \text{ Pa}$$

$$1 \text{ atm} = 101.325 \text{ kPa}$$

$$\text{Throughput: } 1 \text{ W} = 1 \text{ N.m/s} = 7.5 \text{ torr l/s}$$

The composition of the Committee responsible for the formulation of this standard is given in [Annex F](#).

In reporting the result of a test or analysis made in accordance with this standard, if the final value, observed or calculated, is to be rounded off, it shall be done in accordance with IS 2 : 2022 'Rules for rounding off numerical values (*second revision*)'.

Indian Standard

CALIBRATION OF VACUUM GAUGES

PART 1 PRESSURE REDUCTION BY CONTINUOUS FLOW IN THE
PRESSURE RANGE OF 10^{-5} PA TO 10^{-1} PA

(First Revision)

1 SCOPE

1.1 This standard (Part 1) specifies a method for the calibration of vacuum gauges in the pressure range of 10^{-5} Pa to 10^{-1} Pa, whereby a known low pressure is established by the passage of a known flow of gas through a circular orifice of known conductance.

1.1.1 The upper limit of pressure to which the method is applicable is set by the diameter of the orifice in relation to the mean free path of the gas molecules. The lower limit will depend on the type of gauge being calibrated, and on the amount of sorption or desorption of gas within the apparatus.

2 SYMBOLS

The following symbols have been used in the standard:

- p° = Absolute pressure of the gas in the throughput meter;
- p = Calculated pressure of the gas used for calibration in the calibration vessel;
- P_1 = Value of p corrected to relate the calibration of the gauge to a specific temperature, T_0 ;
- Q = Gas throughput, as measured at the temperature, T_Q ;
- S = Net volume rate of flow at the calibration vessel;
- L = Calculated conductance of the orifice;
- S_p = Measured volume rate of flow of the pumping system;
- T_0 = Absolute temperature to which the calibration is to be referred;
- T_c = Absolute temperature of the calibration vessel, including the orifice, during the actual calibration;
- T_Q = Absolute temperature of the gas throughput meter during the actual measurement of the gas throughput in the calibration chamber;
- R = Gas constant;

- M = Molecular mass;
- K_1 = Clausing factor allowing for the thickness of the orifice;
- K_2 = Correction factor allowing for the diameter of the orifice relative to the mean free path of the molecules;
- A_L = Area of the circular orifice;
- r = Radius of the circular orifice;
- \bar{c} = Mean thermal velocity of the molecules ($= \sqrt{\frac{8RT}{\pi M}}$);
- l = Mean free path of the molecules; and
- α = Correction factor for departure from the ideal gas law.

3 APPARATUS**3.1 General Form**

The general form of the apparatus is shown schematically in [Fig. 1](#). The gauges under calibration are connected to an expansion vessel (calibration chamber) into which gas is admitted through a flow measurement device and subsequently pumped away through an orifice.

3.2 Ratio of Cross-Sectional Area of Orifice to Surface Area of Calibration Vessel

The calibration chamber shall be in the form of a sphere or of a right cylinder. The orifice shall be circular, and its cross-sectional area shall be less than 1/1 000 of that of the inner surface of the inscribed sphere of the calibration chamber. The orifice shall be situated in the surface of the inscribed sphere. The volume between the orifice and the pumping system shall be comparable with that of the calibration chamber or even larger.

3.3 Arrangement of Gas Inlet

The gas shall be admitted in such a way that neither the orifice, nor the gauges to be calibrated, nor the gauge connections can be hit by the incoming gas molecules before they have hit the wall at least once.




FIG. 1 SCHEMATIC DIAGRAM OF TYPICAL APPARATUS

3.4 Determination of Orifice Cross-Sectional Area and Conductance

The orifice cross-sectional area A_L shall be determined to an accuracy of ± 0.1 percent. The conductance L is calculated from the area by means of the equation:

$$L = A_L \frac{\bar{c}}{4} K_1 K_2 \dots\dots\dots(1)$$

The correction factors K_1 and K_2 vary with dimensions of orifice (see Annex A).

3.5 Ratio of Rim Thickness to Diameter of Orifice

The rim thickness of the orifice shall be less than 1/50 of its diameter. The influence of the rim

thickness [see equation (1)] in 3.4 is taken into consideration by the clausius factor K_1 , the value of which may be taken from the table given in Annex A. The arrangement for locating the orifice is indicated in Fig. 2. Any device clamping the diaphragm, bearing the orifice, into the calibration chamber shall not protrude into the space occupied by the inscribed spheres above and below the orifice.

3.6 Determination of Temperature of Calibration Vessel

Means shall be provided to measure the temperature T_c of the calibration chamber and orifice, and to keep this temperature constant during the period of the measurements.




FIG. 2 ARRANGEMENT FOR LOCATION OF ORIFICE

3.7 Requirements for Pumping System

The pumping system shall provide for a volume rate of flow of gas through the orifice of at least 10 l/s for the gas used in the calibration. If a vapour pump is used, an adequate baffle or cold trap shall be provided to protect the gauges under calibration against disturbances due to the pump fluid.

3.8 Ratio of Volume Rate of Flow of Pumping System to Conductance of Orifice

The ratio of the measured volume rate of flow, S_p of the pumping system to the conductance L of the orifice shall exceed 50. If then S_p is measured with an accuracy of ± 20 percent, the resulting limit of uncertainty in S is ± 0.4 percent.

3.9 Arrangement of Gauges to be Calibrated

The gauges to be calibrated shall be arranged along the equator of the inscribed sphere of the calibration chamber, with the orifice situated at the pole of this sphere. For the connection of the gauge heads to be calibrated, the calibration vessel shall be provided with connecting pipes, the cross-sections of which correspond to the connecting cross-sections of the gauges, and the lengths of which do not exceed twice the diameter (see 4.3.1).

4 METHOD OF TEST

4.1 Principle

4.1.1 The instrument to be calibrated is connected to a calibration chamber into which gas of accurately known throughput Q is admitted and pumped away through a gas sink of accurately known volume rate of flow S . Under the appropriate conditions the following equation determines the pressure p of the gas in the calibration vessel.

$$p = \frac{Q}{S} \dots\dots\dots (2)$$

4.1.2 Assuming all parts of the calibration equipment to be at the same temperature. The required low pressure p is generated by the method of pressure reduction by continuous flow from a suitable supply of gas at an initial value of pressure sufficiently high to be accurately measurable (see Annex B).

4.2 Preliminary Requirements

4.2.1 Pre-treatment of Calibration Chamber

In order to avoid errors due to other gas sources in the calibration chamber the rate of gas desorption of the apparatus shall be lowered by means of baking to such an extent that the residual gas pressure in the calibration chamber is less than 1/100 of the lowest pressure used during the calibration procedure. This residual gas pressure shall be taken into consideration in correcting the calibration results.

4.2.2 Determination of Volume Rate of Flow

The volume rate of flow S in the calibration chamber is determined by calculation from the conductance L (see 3.4) of the circular orifice and the measured effective volume rate of flow S_p of the pumping system, using the following formula.

$$S = \frac{L}{1+L/S_p} \dots\dots\dots (3)$$

4.2.3 Determination of Effective Volume Rate of Flow of Pumping System

The effective volume rate of flow S_p of the pumping system shall be determined for all conditions and pressure setting representative of the use of the apparatus. One of two methods described below may be used.

4.2.3.1 Measurement of the pressure increase Δp in the calibration chamber, and of the pressure increase Δp_B in the volume between orifice and pumping

system, when additional gas is introduced into the apparatus. The ratio L/S_p is then given by the following equation:

$$\frac{L}{S_p} = \frac{\Delta p B}{\Delta p} \dots\dots\dots (4)$$

The pressure measurement necessitates two similar linear gauges of the same sensitivity, which, however, have not necessarily been calibrated, and which are connected to the calibration chamber and to the volume between orifice and pumping system respectively. In order that these pressure measurements should represent the true existing gas densities unaffected by distortion of the distribution function of the molecules in the volume between orifice and pumping system, this volume shall be about the same as that of the calibration chamber and of similar main dimensions (see 3.2, Fig. 1 and Annex B). The gauges shall be fitted on either side of the orifice on the equators of the two inscribed spheres, representing the volumes above and below the orifice, whose poles coincide at the orifice position.

NOTE — In the case of molecular pumps and diffusion pumps it is sufficient to use the effective volume rate of flow measured for each calibration gas in the required pressure range.

4.2.3.2 Oatley procedure

In this method readings of the pressure p in the calibration chamber are taken by a linear pressure gauge (not necessarily calibrated), at a constant throughput Q of gas, for a series of values of the conductance L . This procedure is repeated for several different values of Q . In order that the value of S_p may be measured with sufficient accuracy, the following conditions shall be met:

- a) The graphs of p when plotted against $1/L$ for a constant value of Q shall not show a significant deviation from a linear relationship; and
- b) If S_p is constant, the straight lines obtained for various values of Q shall intersect at a single point on the axis of $1/L$ given by the equation.

$$\frac{1}{L} + \frac{1}{S_p} = 0 \dots\dots\dots (5)$$

Any deviation from the above requirements gives an estimate of the error involved. The variation of $1/L$ may be determined either by the use of a number of interchangeable orifices or by the use of a single orifice of variable conductance. The ratio of the largest to the smallest area of cross section of the orifice shall be at least 10 : 1. For all orifices used, or for all settings in the case of a variable orifice, the

conditions of 3.5, 4.3.2 and Annex A shall be met. In the case of the Oatley Method, however, an orifice of regular polygonal form may be used if desired.

4.3 Calibration Procedure

4.3.1 Connection of Gauge Heads to be Calibrated (see 3.9)

The gauge heads to be calibrated shall be connected to the calibration chamber by means of connecting pipes, the cross-sections of which correspond to the connecting cross-sections of the gauges and the lengths of which do not exceed twice the diameter. The gauges shall be mounted along the equator of the inscribed sphere of the calibration chamber. The orientation of the gauges to be calibrated shall be well defined. Alternatively, it shall be ensured that their orientation does not significantly affect the calibration. Nude systems shall be connected in such a way that they are immersed as far as possible. If more than one nude gauge is attached to the calibration chamber, only one gauge shall be operated at any time, unless it is certain that a mutual influence is not possible.

4.3.2 Limitation of Effects due to Pumping and Gas Desorption of the Gauges to be Calibrated

In order to keep the errors resulting from the gauges to be calibrated acting as pumps or sources of gas, sufficiently small, the following conditions shall be met:

- a) The volume rate of flow due to pumping by the connected gauges, or their gas desorption rate expressed in the same units, shall not exceed 1/100 of the volume rate of flow S at the calibration chamber; and
- b) The volume rate of flow S at the calibration chamber shall be at least 10 l/s for the gases used in the calibration procedure, even if compliance with condition (a) shall allow a lower value.

4.3.2.1 If the above conditions cannot be met, the results shall be corrected by means of the indications of a suitable auxiliary gauge. The conditions (a) and (b) above are especially important in the case of hot-cathode ionization gauges operating with gases, such as hydrogen, methane, oxygen and air.

4.3.2.2 In certain cases it may be necessary to check for possible anomalous ion current caused by desorption of ions from the grid. For this purpose the ion current characteristic shall be recorded in order to ensure that the calibration relates to conditions where this characteristic is linear.

4.3.3 Operation of Gauges to be Calibrated

All the operating requirements specified by the manufacturer of the gauge and its associated parts shall be carefully followed. For gauges which have several indicated ranges the exact relations between them shall be the subject of a separate test.

4.3.4 Determination of Gas Throughput Q

An effective volume rate of flow S at the calibration chamber shall be 10 l/s or more. The corresponding range of the throughput measurement is therefore between 10^{-7} Pa.m³/s and a value above 10^{-3} Pa.m³/s. This measurement shall have a typical uncertainty of ± 0.5 percent which shall be confirmed by periodic checks. The temperature T_Q of the gas during the measurement of throughput shall be known to within ± 0.1 percent, corresponding to ± 0.3 K at 300 K.

4.3.5 Other Aspects of Calibration Technique

4.3.5.1 Attainment of equilibrium

In order that errors due to adsorption and desorption effects are made negligibly small, it is essential to wait for a sufficient time for all indications and readings to reach constancy within appropriate limits of uncertainty.

4.3.5.2 Temperature requirements

The temperature of the gas throughput meter, T_Q and of the calibration chamber T_c , shall not deviate from the reference temperature T_0 by more than ± 10 °C. The reference temperature shall lie in the range 20 °C to 25 °C but preferably at 23 °C. The temperature of the calibration chamber and orifice T_c and that of the throughput meter T_Q shall be kept constant during the period of measurements, this can be advantageously accomplished by means of a liquid closed-loop system. The values of T_c and T_Q shall be measured to within about ± 0.1 percent, corresponding to ± 0.3 K at 300 K.

4.4 Choice of Calibration Gas

4.4.1 Purity

For the calibration procedure pure gases should preferably be used. In each case the maximum content of any additional components present shall be known, so that their influence upon the calibration result can be evaluated. It shall be ensured that such additional components do not cause errors exceeding 0.1 percent.

4.4.2 Drying

The gases shall be dried to a dewpoint which corresponds to a water vapour pressure of less than one part in 10^5 parts of gas by volume. This corresponds to a pressure p^0 of 1 atm at a dewpoint of approximately -60 °C.

4.4.3 Preferred Gases

For the calibration procedure, argon and nitrogen are preferred unless otherwise specified. In some technical applications, however, air is frequently used. As the molecular masses of the main components of air are very similar, it is permissible to take air as a gas of uniform composition as long as it is ascertained that all components are about equally well pumped by the pumping system. Ion pumps shall not be used in the latter case. An effective molecular mass for air shall be used which is given by:

$$M_{\text{eff}} = \left(\sum_i X_i \sqrt{M_i} \right)^2 = 28.9 \dots \dots \dots (6)$$

where X_i is the mole fraction of the component of molecular mass M_i . However, the calibration may correspond to a changed composition of air depending on the characteristics of the variable leak valve. If this valve exhibits molecular flow, the composition of the air in the calibration vessel is unchanged:

$X_i = 78.1$ percent, 21.0 percent and 0.9 percent for nitrogen, oxygen and argon respectively.

If the valve exhibits purely viscous flow, the composition of the air in the calibration vessel is:

$X_i = 76.8$ percent, 22.1 percent and 1.1 percent for nitrogen, oxygen and argon respectively.

4.4.3.1 Possible effects in calibration with gas mixtures are given in [Annex D](#).

4.5 Interpretation of Measurements

4.5.1 Effects of Non-uniformity of Temperature

In the ideal case where all parts of the equipment are at the same temperature, the pressure activating the gauge is given in terms of the throughput Q and the volume rate of flow of the orifice S by equation (2) (see 4.1). If, with the device for measuring the gas throughput Q at the absolute temperature T_Q and with the calibration chamber and therefore also the orifice at the absolute temperature T_c the calibration has to be related to a reference absolute temperature T_0 , then instead of equation (2), the following equations apply:

- a) For a gauge that responds to the pressure of the gas:

$$p_1 = \frac{Q}{S T_0} \frac{\sqrt{T_0 T_c}}{T_Q} \dots \dots \dots (7)$$

- b) For a gauge that responds to the density of the gas:

$$p_1 = \frac{Q}{S T_0} \frac{T_0}{T_Q} \frac{\sqrt{T_0}}{\sqrt{T_c}} \dots \dots \dots (8)$$

Equation (7) and Equation (8) give the pressures p_1 which should be used in determining the sensitivity of the gauge to be calibrated, if the calibration is to be related to the absolute temperature T_0 .

4.5.2 Corrections for Departures from the Ideal Gas Law

In the case of real gases a correction may be necessary to account for departure from the ideal gas law over the range from the initial pressure p^0 down to the very small calibration pressure p . If α is the ratio of the pV -values at the expansion pressure $p \rightarrow 0$ and the pressure p^0 at the same temperature T_Q , then the true pressure in the calibration chamber is given by the following formula:

$$p_{\text{true}} = p \alpha, \text{ with } \alpha = \frac{pV(p^0 \rightarrow 0, T_Q)}{pV(p^0, T_Q)} \dots \dots \dots (9)$$

Here α refers to the gas at temperature T_Q . If the initial starting pressure p^0 is approximately 1 atm, then for the following gases the correction is less than ± 0.1 percent and may therefore be neglected:

Helium, neon, argon, hydrogen, oxygen, nitrogen, air free from carbon dioxide, and carbon monoxide.

4.5.2.1 Correction factors for various gases are given in [Annex E](#).

4.5.3 Limits of Uncertainty

Careful attention shall be paid to all factors influencing the accuracy of measurements and consideration of the corresponding limits of uncertainty (see [Annex C](#)).

5 TEST REPORT

As far as applicable to the various types of gauges, the test report shall contain the following information:

- a) The date of calibration;
- b) The manufacturer of the gauge, the model and serial number;
- c) The position and geometry of the connecting flange during calibration, and the orientation of the gauge;
- d) The geometry of important parts of the gauge and the material, as far as this is disclosed by the manufacturer, and if relevant to the purpose of the calibration;
- e) The temperature of the calibration chamber, of the gas throughput meter and of the ambient space;
- f) The reference temperature T_0 for which the calibration is valid;
- g) The temperature and duration of baking, de-gassing and ageing;
- h) The previous history of the gauge as far as this is essential for the calibration;
- j) If the calibration of the gauge head is carried out separately from the supply unit, all currents, voltages and the other relevant quantities recorded during the calibration procedure;
- k) If the calibration of the gauge head is carried out together with the supply unit, all values set at the supply unit, the ranges used and the mains voltage;
- m) The calibration curve which shows the sensitivity related to the reference temperature as a function of pressure;
- n) The effective volume rate of flow at the calibration chamber during calibration;
- p) The composition of the gas used for calibration;
- q) The details of the calibration method used, including reference to any appropriate national or international standard;
- r) The limits of uncertainty of the calibration;
- s) Any noted tendency for the gauges to act as sources or sinks of gas; and
- t) The limits of uncertainty within which the calibration system is maintained in equilibrium during calibration.

ANNEX A

(Clauses 3.4, 3.5, 4.2.3.2 and C-2)

CORRECTIONS INVOLVED IN CALCULATION OF CONDUCTANCE OF ORIFICE

A-1 GENERAL

A-1.1 This annex gives further information on the values of the two correction factors K_1 and K_2 which may be involved in the calculation of the conductance of the orifice by means of the formula given in 3.4.

A-2 CLAUSING FACTOR

A-2.1 The clausing factor allowing for thickness of orifice is given below:

<i>Sl No.</i>	<i>Thickness Radius</i>	<i>Clausing Factor, K_1</i>
(1)	(2)	(3)
i)	0.001	0.999 5
ii)	0.002	0.999 0
iii)	0.003	0.998 5
iv)	0.004	0.998 0
v)	0.005	0.997 5
vi)	0.006	0.997 0
vii)	0.007	0.996 5
viii)	0.008	0.996 0
ix)	0.009	0.995 5
x)	0.010	0.995 0
xi)	0.011	0.994 5
xii)	0.012	0.994 0
xiii)	0.013	0.993 5
xiv)	0.014	0.993 0

<i>Sl No.</i>	<i>Thickness Radius</i>	<i>Clausing Factor, K_1</i>
(1)	(2)	(3)
xv)	0.015	0.992 6
xvi)	0.016	0.992 1
xvii)	0.017	0.991 6
xviii)	0.018	0.991 1
xix)	0.019	0.990 6
xx)	0.020	0.990 1

A-3 UNCERTAINTY LIMITS

A-3.1 Relation between Orifice Dimensions and Uncertainty Limits of Calibration Pressure

At the higher end of the pressure range used in the calibration, the mean-free path l of the molecules may no longer be very large compared with the radius r of the orifice. This resulting error may be corrected by the factor K_2 , where

$$K_2 = 1 + \frac{r}{4l} \dots\dots\dots (10)$$

It is desirable that K_2 should not exceed 1.03. If then one assumes an uncertainty in this correction of ± 10 percent, the maximum uncertainty resulting from the determination of the conductance will not exceed ± 0.3 percent. This implies that, in the case of argon at a pressure of 10^{-1} Pa, the radius r of the orifice shall not exceed 5 mm.

ANNEX B

(Clauses [4.1.2](#) and [4.2.3.1](#))

PREREQUISITES FOR THE APPLICABILITY OF EQUATION (2)

B-1 PREREQUISITES

B-1.1 Equation (2) holds strictly only in the case of an undistorted distribution function of the gas molecules in the calibration chamber. A distortion of the distribution may arrive from either the gas inlet or the gas outlet regions, and this may restrict the applicability of equation (2) in three ways:

- a) The gas pressure in the calibration vessel may no longer be well defined in the sense of statistical mechanics;
- b) The gas density may no longer be uniform; and
- c) The calculation of the conductance of the orifice from its geometry may require a gas-kinetic correction.

B-1.1.1 For these reasons the chamber, gas inlet and orifice shall be so dimensioned and arranged that the distortion becomes sufficiently small. In a spherical calibration chamber the distortion caused by gas inlet and gas sink may easily be evaluated. In the case of molecules characterized by an undistorted

velocity distribution function an equal number of molecules impinges on all unit surface elements of the inner surface of the hollow sphere. Also, the molecules emitted by any surface element are distributed evenly over the inner surface, provided Lambert's Law (Cosine Law) applies. It follows, as an adequate approximation in the case of molecular flow that the proportionate disturbance caused by a circular hole, such as the outlet orifice does not exceed the ratio of the cross-sectional area of the orifice to the inner surface area of the sphere. If the vessel is not a hollow sphere, the greatest inscribed sphere shall be used as a basis for the dimensioning of the system. The geometry of the gas inlet shall take account of the fact that the directions of the entering molecules usually do not comply with the cosine law. This is guaranteed, to a great extent, after the first impact on a wall.

B-1.1.2 Similarly, the volume between the orifice and the pumping system shall be as large as the calibration chamber to ensure an undisturbed velocity distribution function in the volume which is responsible for back diffusion of the calibration gas through the orifice.

ANNEX C

(Clause 4.5.3) DISCUSSION
OF ERRORS

C-1 GENERAL EXPRESSION FOR THE ERROR

According to equations (7) and (8) the pressure p_1 is given by:

$$p_1 = \frac{Q}{sT_0} \frac{\sqrt{T_0 T_c}}{T_Q} \text{ or } p_1 = \frac{Q}{sT_0} \frac{T_0}{T_Q} \sqrt{\frac{T_0}{T_c}}$$

Using equation (7) and taking into consideration the equations (3) and (1) it follows that:

$$p_1 = \frac{Q}{A_L \frac{c}{4} K_1 K_2 \left(\frac{1}{1+L/S_P} \right)} \frac{\sqrt{T_0 T_c}}{T_Q} \dots\dots\dots (11)$$

and with a corresponding expression when equation (8) is used. From this the relative error is given by:

$$\frac{\delta p_1}{p_1} = \frac{\delta Q}{Q} + \frac{\delta A_L}{A_L} + \frac{\delta K_2}{K_2} + \frac{\delta \beta}{\beta} + \frac{1}{2} \frac{\delta T_c}{T_c} + \frac{\delta T_Q}{T_Q} \dots\dots\dots (12)$$

In this equation, $\beta = \frac{1}{1+L/S_P}$

Because of this smallness the relative error of the clausung correction has not been taken into consideration. A corresponding expression may be obtained for the error in the case of the Oatley procedure, for determining the effective volume rate of flow of the pumping system (see 4.2.3).

C-2 MAGNITUDE OF TOTAL ERROR

The total relative error [see equation (12)] is numerically made up in the following way:

- a) $\frac{\delta Q}{Q}$ shall be of the order of ± 0.5 percent and

depends on the type of throughput meter used. The relative error of the throughput measurement shall be determined under the prevailing conditions for each gas and pressure setting;

- b) $\frac{\delta A_L}{A_L} = 0.1$ percent (see 3.4);
- c) $\frac{\delta K_2}{K_2} = 0.3$ percent (see Annex A);
- d) $\frac{\delta \beta}{\beta} = 0.4$ percent (see 3.8 and 4.2.3);
- e) $\frac{\delta T_c}{T_c} = 0.1$ percent (see 3.6); and
- f) $\frac{\delta T_Q}{T_Q} = 0.1$ percent (see 4.3.4).

The total relative error therefore is:

$$\frac{\delta p_1}{p_1} \leq 1 \text{ percent} + \frac{\delta Q}{Q} \dots\dots\dots (13)$$

The relative error $\frac{\delta Q}{Q}$ shall be determined afresh for each measurement.

C-3 DIMENSIONING OF THE APPARATUS

Dimensions for the various components of the apparatus are not given. However, the dimensions, particularly of the gas throughput meter, shall be such that the expression $\delta Q/Q$ becomes a minimum. The most suitable dimensions depend essentially upon the properties of the components and their stability in time, as well as upon a number of other requirements such as the pressure range of the calibration, the time required for one calibration.

ANNEX D

(Clause [4.4.3.1](#))

ASPECTS OF CALIBRATION USING GAS MIXTURES

D-1 GENERAL

Complications may arise from the use of gas mixtures due to the fact that molecular flow through the orifice causes the components to become separated if the molecular masses are different. Similarly, separation may occur during the flow from the gas throughput meter, through the variable leak valve, into the calibration vessel. Such errors caused by the variable leak valve can be eliminated by using an uncalibrated fixed leak which exhibits either:

- a) Purely viscous flow; or
- b) Purely molecular flow, in the pressure range used.

It should also be noted that the effective volume rates of flow of the pumps and the mean-free-paths of the molecules flow are different for the various components.

D-2 USE OF GAS MIXTURES WITH A VISCOUS FLOW LEAK

D-2.1 When using a fixed leak exhibiting purely viscous flow for the expansion, instead of the equations (7) and (8) the following expressions shall be used:

- a) For gauges responding to pressure:

$$p_1 = \frac{Q\sqrt{T_0 T_C}}{T_Q} \sum_i \frac{X_i}{S_i T_0} \dots\dots\dots (14)$$

- b) For gauges responding to density:

$$p_1 = \frac{QT_0}{T_Q} \sqrt{\frac{T_0}{T_C}} \sum_i \frac{X_i}{S_i T_0} \dots\dots\dots (15)$$

where

X_i = The mole fraction of the gas component i with $\sum_i X_i = 1$; and

S_i = The corresponding rate of flow.

The calibration now no longer refers to the original gas mixture having the molecular fractions X_i , but to a mixture in the expansion volume with the new molecular fractions:

$$X_{ii} = \frac{X_i \sqrt{M_i}}{\sum_i X_i \sqrt{M_i}}$$

D-3 USE OF GAS MIXTURES WITH A MOLECULAR FLOW LEAK

When using a fixed leak exhibiting purely molecular flow for the expansion, the equation (7) and equation (8) remain valid if the effective molecular mass:

$$M_{\text{eff}} = \left(\sum_i X_i \sqrt{M_i} \right)^2$$

is substituted for S in the expression, the calibration referring to the original gas mixture.

ANNEX E

(Clause [4.5.2.1](#))

TABLE OF CORRECTIONS ARISING FROM DEPARTURES FROM THE IDEAL GAS LAW

E-1 CORRECTION FACTOR

The following table lists the correction factor α given by the expression:

$$\alpha = \frac{pV(p \rightarrow 0, 25^\circ \text{C})}{pV(1 \text{ atm}, 25^\circ \text{C})}$$

corresponding to the gas temperature 25 °C, by which the pressure p estimated from equations (7) and (8) shall be multiplied to obtain the true pressure,

$$p_{\text{true}} = p \alpha$$

He	0.999 5
H ₂ , D ₂	0.999 5
NH ₃	1.012 0
CH ₄	1.001 9
C ₃ H ₆	1.014 4
Ne	0.999 6
O ₂	1.000 6
N ₂ O	1.005 0
C ₂ H ₆	1.007 8
Ar	1.000 7
SF ₆	1.011 7
CO	1.000 4
C ₂ H ₄	1.005 4
Kr	1.002 2
N ₂	1.000 2
CO ₂	1.005 5
C ₂ H ₂	1.006 9
Xe	1.005 5
Air free from CO ₂	1.000 4
CF ₄	1.003 8
C ₃ H ₈	1.015 4

The figures given in this table refer to the case where the gas is expanded at room temperature starting from about 1 atm.

ANNEX F

(Foreword)

COMMITTEE COMPOSITION

Chemical Engineering Plants and Related Equipment Sectional Committee, MED 17

<i>Organization</i>	<i>Representative(s)</i>
CSIR - Indian Institute of Petroleum, Dehradun	DR MRITUNJAY KUMAR SHUKLA (Chairperson)
Auma India Private Limited, Bengaluru	SHRI YASHWANT M. JANNU
Bharat Heavy Electrical Limited, New Delhi	SHRI Y. SRINIVASA RAO SHRI ABHISHEK KUMAR PANDEY (<i>Alternate I</i>) SHRI RAJESH RANJAN (<i>Alternate II</i>) SHRI SUBHASHISH GUPTA (<i>Alternate III</i>)
Bharat Petroleum Corporation Limited Corporate Research & Development Centre, Greater Noida	MS ISHA KHULLAR SHRI VINOD KUMAR (<i>Alternate</i>)
Blast Carboblocks Private Limited, Mumbai	SHRI DHAWAL SAXENA
Chemtrols Industries Private Limited, New Delhi	SHRI P. KRISHNA KUMAR
Confederation of Indian Industry, New Delhi	SHRI NANDAKUMAR KALATH SHRI ABILASH UTTAM (<i>Alternate</i>)
Directorate General Factory Advice Service and Labour Institutes, Mumbai	SHRI TANOJ CHANDAN SHRI KUNAL SHARMA (<i>Alternate</i>)
Engineers India Limited, Gurugram	SHRI HASMUKH K. PARMAR SHRI MRAGANG SHEAKHAR (<i>Alternate</i>)
Fab-Tech Works and Constructions Private Limited, Mumbai	SHRI AASHISH JAYPRAKASH LAKHANI SHRI PRADEEP GAWATE (<i>Alternate</i>)
GMM Pfaudler Limited, Anand	SHRI DHIRAN PANCHAL SHRI SATVIK PATEL (<i>Alternate</i>)
Kejriwal Casting Limited, Kolkata	SHRI SANDEEP KEJRIWAL SHRI SABARNA ROY (<i>Alternate</i>)
L&T Valves, Chennai	SHRI ROHIT SHARMA SHRI SURIYANARAYANAN (<i>Alternate</i>)
MECON Limited, Ranchi	SHRI YOGENDRA KUMAR SINGH SHRI ARVIND BHUSHAN (<i>Alternate</i>)
Nuclear Power Corporation of India Limited, Mumbai	SHRI CHANDRAKANT RAJARAM KAKADE SHRI ARUNAVA SINHA
Project and Development India Limited, Noida	SHRI SANJIV KUMAR MISHRA SHRI RAJEEV RANJAN KUMAR (<i>Alternate</i>)
Tata Consulting Engineers Limited, Navi Mumbai	SHRI SHIVNARAYAN PAREEK SHRI SHIREESH S. SWAMI (<i>Alternate</i>)
BIS Directorate General	SHRI K. VENKATESWARA RAO, SCIENTIST 'F'/SENIOR DIRECTOR AND HEAD (MECHANICAL ENGINEERING) [REPRESENTING DIRECTOR GENERAL (<i>Ex-officio</i>)]

Member Secretary
MS NEHA THAKUR
SCIENTIST 'C'/DEPUTY DIRECTOR
(MECHANICAL ENGINEERING), BIS

Bureau of Indian Standards

BIS is a statutory institution established under the *Bureau of Indian Standards Act, 2016* to promote harmonious development of the activities of standardization, marking and quality certification of goods and attending to connected matters in the country.

Copyright

BIS has the copyright of all its publications. No part of these publications may be reproduced in any form without the prior permission in writing of BIS. This does not preclude the free use, in the course of implementing the standard, of necessary details, such as symbols and sizes, type or grade designations. Enquiries relating to copyright be addressed to the Head (Publication & Sales), BIS.

Review of Indian Standards

Amendments are issued to standards as the need arises on the basis of comments. Standards are also reviewed periodically; a standard along with amendments is reaffirmed when such review indicates that no changes are needed; if the review indicates that changes are needed, it is taken up for revision. Users of Indian Standards should ascertain that they are in possession of the latest amendments or edition by referring to the website-www.bis.gov.in or www.standardsbis.in.

This Indian Standard has been developed from Doc No.: MED 17 (23611).

Amendments Issued Since Publication

Amend No.	Date of Issue	Text Affected

BUREAU OF INDIAN STANDARDS

Headquarters:

Manak Bhavan, 9 Bahadur Shah Zafar Marg, New Delhi 110002

Telephones: 2323 0131, 2323 3375, 2323 9402

Website: www.bis.gov.in

Regional Offices:

Central : 601/A, Konnectus Tower -1, 6th Floor,
DMRC Building, Bhavbhuti Marg, New
Delhi 110002

Telephones

{ 2323 7617

Eastern : 8th Floor, Plot No 7/7 & 7/8, CP Block, Sector V,
Salt Lake, Kolkata, West Bengal 700091

{ 2367 0012
{ 2320 9474

Northern : Plot No. 4-A, Sector 27-B, Madhya Marg,
Chandigarh 160019

{ 265 9930

Southern : C.I.T. Campus, IV Cross Road, Taramani, Chennai 600113

{ 2254 1442
{ 2254 1216

Western : 5th Floor/MTNL CETTM, Technology Street, Hiranandani Gardens,
Powai, Mumbai 400076

{ 2570 0030
{ 2570 2715

Branches : AHMEDABAD, BENGALURU, BHOPAL, BHUBANESHWAR, CHANDIGARH, CHENNAI, COIMBATORE, DEHRADUN, DELHI, FARIDABAD, GHAZIABAD, GUWAHATI, HARYANA (CHANDIGARH), HUBLI, HYDERABAD, JAIPUR, JAMMU, JAMSHEDPUR, KOCHI, KOLKATA, LUCKNOW, MADURAI, MUMBAI, NAGPUR, NOIDA, PARWANOO, PATNA, PUNE, RAIPUR, RAJKOT, SURAT, VIJAYAWADA.