

सीबीएम मूल्यांकन के लिए प्रयोगशाला
उपकरण — रीति संहिता
भाग 2 मार्श फनल और अंशांकित कप

Lab Instruments for CBM
Evaluation — Code of Practice
Part 2 Marsh Funnel and Graduated Cup

ICS 73.020

© BIS 2024



भारतीय मानक ब्यूरो
BUREAU OF INDIAN STANDARDS
मानक भवन, 9 बहादुर शाह ज़फर मार्ग, नई दिल्ली - 110002
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG
NEW DELHI - 110002
www.bis.gov.in www.standardsbis.in

FOREWORD

This Indian Standard (Part 2) was adopted by the Bureau of Indian Standards, after the draft finalized by the Method and Equipments for Underground Coal Gasification and Coal Bed Methane Sectional Committee had been approved by the Mechanical Engineering Divisional Council.

Coal bed methane is the form of natural gas that is adsorbed into the solid matrix of coal. It is different from the conventional gas reservoirs as the methane is stored within the coal seams through the process of adsorption. The natural fractures in the coal seams (known as cleats) are responsible for the flow behaviour and provide the major channels for gas flow.

Several laboratory studies are carried out for CBM exploration/extraction as well as at the time of CBM operations. The laboratory studies include with various geological and geochemical data collection, data analysis, quality check and quality control of operation fluid while drilling, hydro-fracturing etc. The lab studies are essential to evaluate the gas reserve as well as for strategy finalization for exploration methodologies. One of the most challenging tasks in CBM evaluation are effluent (produced water) handling.

The code of practices for lab instruments for CBM evaluation is in four parts. This standard (Part 2) covers the marsh funnel and graduated cup. Other parts in this series under the general title are as follows:

Part 1 Mud balance

Part 3 Turbidity meter

Part 4 Rotational viscometer

The composition of the Committee responsible for the formulation of this standard is given in [Annex B](#).

For the purpose of deciding whether a particular requirement of this standard is complied with, the final value, observed or calculated, expressing the result of a test or analysis, shall be rounded off in accordance with IS 2 : 2022 'Rules for rounding off numerical values (*second revision*)'. The number of significant places retained in the rounded-off value should be the same as that of the specified value in this standard.

Indian Standard

LAB INSTRUMENTS FOR CBM EVALUATION — CODE OF PRACTICE

PART 2 MARSH FUNNEL AND GRADUATED CUP**1 SCOPE**

This standard covers a marsh funnel which is a marsh cone with a particular orifice and diameter and with a fixed capacity. The graduated cup is also a fixed volume cup for measurement of funnel viscosity. It is a simple device for measurement of indicating viscosity of drilling fluid/or any viscous liquid for quality check, laboratory analysis and design purpose. Marsh funnel is a useful and utility device in drilling operation for quick quality analysis of drilling fluid. Funnel viscosity is the indicating viscosity of a fluid and it is measured by observing the time it takes a known volume of liquid to flow from a cone through a short tube.

2 MARSH FUNNEL AND GRADUATED CUP**2.1 Design and Requirement****2.1.1 Marsh Funnel**

- a) Material — It should be made of rugged break/crack resistant plastic or equivalent material to resist temperature change deformation;
- b) Funnel diameter — 6" (152 mm) at the top;
- c) Length — 12" (305 mm);
- d) Capacity of the cone or funnel to bottom of screen — 1500 cc; and
- e) Plastic handle on cone/funnel to protect hands during testing.

2.1.2 Screen

- a) A 12 mesh half circle screen designed for removing any foreign material and drill cutting; and
- b) Screens should be fixed at a level 0.748" (19 mm) below edge top of the funnel.

2.1.3 Orifice

- a) Material — Copper or brass;
- b) Length — 2.0" (50.8 mm); and
- c) Diameter — (Inner diameter) 0.185" (4.7 mm).

2.1.4 Measuring Cup

Graduated in cc/ounce. Capacity 1 000 ml with moulded mark at 350 ml and at 946 ml.

2.1.5 Material

It should be made of rugged break/crack resistant plastic or equivalent material to resist temperature change deformation.

2.1.6 Accuracy

26 s ± 0.5 s at a temperature of 70 °F ± 5 °F (21°C ± 3 °C) that is, time required to drain one quart (946 ml) by measuring the viscosity of fresh water.

2.1.7 Other Accessories

- a) Digital stop watch for time recording;
- b) Digital thermometer (Fahrenheit and centigrade) as (0 °C to 105 °C)/(32 °F to 220 °F); and
- c) T-handle reamer 4.7 mm.

3 CALIBRATION

The device should frequently be checked with fresh water. Fresh water should give the reading of 26 s ± 0.5 s at a temperature of 70 °F ± 5 °F (21°C ± 3°C) that is, time required to drain one quart (946 ml).

4 INSTALLATION AND TESTING

- a) Funnel viscosity (that is, indication viscosity) of any fluid is measured by using marsh funnel and graduated cup;
- b) Cover the funnel orifice with a finger and pour freshly sampled drilling fluid/sample fluid through the screen up to the bottom of the screen;
- c) Remove finger and start the stopwatch simultaneously. Record the time of filling of measuring fluid to fill to the 946 ml that is to the mark of the cup; and
- d) Record the temperature of the fluid and report funnel viscosity of the fluid as nearest second of time in that temperature.

5 DOCUMENTATION

- a) Past track record of at least last two years for marsh funnel supplied in oil industry for example purchase orders, inspection release notes, user feedback with their communication details etc shall be provided;
- b) Manufacturer's data and descriptive literature for the equipment and materials of construction by ASTM reference and grade, coating(s), etc specifications;
- c) Calibration and test report of manufacturer shall be furnished;
- d) Installation and operation manual in English language shall be provided along with the supply; and
- e) Warranty/Performance guarantee certificate shall be furnished.

ANNEX A

(Normative)

MARSH FUNNEL AND GRADUATED CUP— DATA SHEET

<i>Sl No.</i>	<i>Description</i>	<i>Details</i>
(1)	(2)	(3)
i)	Marsh cone (dimension)	-
ii)	Marsh funnel capacity	1 500 CC
iii)	Screen (mesh)	Half circle 12 mesh screen
iv)	Orifice diameter size)	0.185" (4.7mm), inner diameter
v)	Graduated cup capacity (ml) and material	1 000 ml with moulded mark at 350 ml and at 946 ml
vi)	Digital stop watch	-
vii)	Digital thermometer (Fahrenheit and centigrade)	-
viii)	Accuracy	26 s \pm 0.5 s at a temperature of 70 °F \pm 5°F (21°C \pm 3°C)
ix)	Conformance to codes and standards (if applicable)	-

ANNEX B

(Foreword)

COMMITTEE COMPOSITION

Method and Equipments for Underground Coal Gasification and Coal Bed Methane Sectional Committee, MED 37

<i>Organization</i>	<i>Representative(s)</i>
Oil and Natural Gas Corporation Limited, New Delhi	SHRI UDAY PASWAN (Chairperson)
Atlas Copco Construction and Mining Sales, Pune	SHRI ANIMESH NANDY
Bharat Heavy Electrical Limited, New Delhi	SHRI TIRUPATHI NAIDU CHINTALA
Bharat Heavy Electricals Limited, Project Engineering Management, Noida	SHRI RAJESH RANJAN SHRI SAUMEN KUMAR BHAUMIK (<i>Alternate I</i>) SHRI PRADEEP KUMAR SHARMA (<i>Alternate II</i>)
Central Electricity Authority, New Delhi	SHRI SUNIT GUPTA SHRI ASIF IQBAL DEPUTY (<i>Alternate</i>)
Central Mine Planning and Design Institute Limited, Ranchi	DR AKHILESH SINGH
CSIR - Central Institute for Mining and Fuel Research, Dhanbad	DR DEBADUTTA MOHANTY SHRI JAYWARDHAN KUMAR (<i>Alternate</i>)
CSIR - Central Mechanical Engineering Research Institute, Durgapur	DR MALAY KUMAR KARMAKAR DR CHANCHAL LOHA (<i>Alternate</i>)
Directorate General of Hydrocarbons, Noida	MS AARTI GUPTA SHRI TRILOK NATH (<i>Alternate</i>)
Directorate General of Mines Safety, Dhanbad	SHRI SAIFULLAH ANSARI SHRI A. RAJESHWAR RAO (<i>Alternate</i>)
Essar Oil and Gas Exploration and Production Limited, Durgapur	SHRI VINEET SINGHAL SHRI VIKRAM A. GODAY (<i>Alternate</i>)
GAIL (India) Limited, New Delhi	SHRI RAJESH BAGARIA SHRI A. K. PORWAL (<i>Alternate</i>)
Great Eastern Energy Corporation Limited, Asansol	SHRI ANOOP GUPTA SHRI PRIYARANJAN PATRA (<i>Alternate</i>)
Indian Institute of Technology (ISM), Dhanbad	SHRI R. M. BHATTACHARJEE SHRI D. P. MISHRA (<i>Alternate</i>)
Oil and Natural Gas Corporation Limited, New Delhi	SHRI A. K. PASWAN SHRI SHAKEEL AHMED (<i>Alternate</i>)
In Personal Capacity (<i>Flat No. 3052, Prestige Shantiniketa, Whitefield main road, Bengaluru</i>)	SHRI R. K. SHARMA
In Personal Capacity (<i>D-24, Amar Colony, New Delhi</i>)	SHRI RUDRA PRATAP SINGH

Organization

BIS Directorate General

Representative(s)

SHRI K. V. RAO, SCIENTIST 'F'/SENIOR DIRECTOR AND
HEAD (MECHANICAL ENGINEERING)[REPRESENTING
GENERAL (*Ex-officio*)]

Member Secretary

SHRI AMAN DHANAWAT
SCIENTIST 'C'/DEPUTY DIRECTOR
(MECHANICAL ENGINEERING), BIS

Bureau of Indian Standards

BIS is a statutory institution established under the *Bureau of Indian Standards Act, 2016* to promote harmonious development of the activities of standardization, marking and quality certification of goods and attending to connected matters in the country.

Copyright

BIS has the copyright of all its publications. No part of these publications may be reproduced in any form without the prior permission in writing of BIS. This does not preclude the free use, in the course of implementing the standard, of necessary details, such as symbols and sizes, type or grade designations. Enquiries relating to copyright be addressed to the Head (Publication & Sales), BIS.

Review of Indian Standards

Amendments are issued to standards as the need arises on the basis of comments. Standards are also reviewed periodically; a standard along with amendments is reaffirmed when such review indicates that no changes are needed; if the review indicates that changes are needed, it is taken up for revision. Users of Indian Standards should ascertain that they are in possession of the latest amendments or edition by referring to the website-www.bis.gov.in or www.standardsbis.in.

This Indian Standard has been developed from Doc No.: MED 37 (20187).

Amendments Issued Since Publication

Amend No.	Date of Issue	Text Affected

BUREAU OF INDIAN STANDARDS

Headquarters:

Manak Bhavan, 9 Bahadur Shah Zafar Marg, New Delhi 110002

Telephones: 2323 0131, 2323 3375, 2323 9402

Website: www.bis.gov.in

Regional Offices:

Central : 601/A, Konnectus Tower -1, 6th Floor,
DMRC Building, Bhavbhuti Marg, New
Delhi 110002

Telephones

{ 2323 7617

Eastern : 8th Floor, Plot No 7/7 & 7/8, CP Block, Sector V,
Salt Lake, Kolkata, West Bengal 700091

{ 2367 0012
{ 2320 9474

Northern : Plot No. 4-A, Sector 27-B, Madhya Marg,
Chandigarh 160019

{ 265 9930

Southern : C.I.T. Campus, IV Cross Road, Taramani, Chennai 600113

{ 2254 1442
{ 2254 1216

Western : 5th Floor/MTNL CETTM, Technology Street, Hiranandani Gardens,
Powai, Mumbai 400076

{ 2570 0030
{ 2570 2715

Branches : AHMEDABAD, BENGALURU, BHOPAL, BHUBANESHWAR, CHANDIGARH, CHENNAI, COIMBATORE, DEHRADUN, DELHI, FARIDABAD, GHAZIABAD, GUWAHATI, HARYANA (CHANDIGARH), HUBLI, HYDERABAD, JAIPUR, JAMMU, JAMSHEDPUR, KOCHI, KOLKATA, LUCKNOW, MADURAI, MUMBAI, NAGPUR, NOIDA, PARWANOO, PATNA, PUNE, RAIPUR, RAJKOT, SURAT, VIJAYAWADA.