

रक्त संक्रमण के लिए फिल्टर और फिल्टर
चैम्बर — विशिष्टि
(पहला पुनरीक्षण)

Filter and Filter Chamber for Blood
Transfusion — Specification

(First Revision)

ICS 11.040.20

© BIS 2024



भारतीय मानक ब्यूरो
BUREAU OF INDIAN STANDARDS
मानक भवन, 9 बहादुर शाह ज़फर मार्ग, नई दिल्ली - 110002
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG
NEW DELHI - 110002

www.bis.gov.in www.standardsbis.in

FOREWORD

This Indian Standard (First Revision) was adopted by the Bureau of Indian Standards after the draft finalized by the Hospital Equipment and Surgical Disposable Products Sectional Committee had been approved by the Medical Equipment and Hospital Planning Division Council.

This standard was first published in 1967 with the title 'Specification for Filter and filter chamber for blood transfusion'. This revision has been brought out to align the cross-references to the latest editions.

The composition of the Committee responsible for formulation of this standard is given in [Annex A](#).

For the purpose of deciding whether a particular requirement of this standard is complied with the final value, observed or calculated, expressing the result of a test or analysis shall be rounded off in accordance with IS 2 : 2022 'Rules for rounding off numerical values (*second revision*)'. The number of significant places retained in the rounded off value should be same as that of the specified value in this standard.

*Indian Standard***FILTER AND FILTER CHAMBER FOR BLOOD
TRANSFUSION — SPECIFICATION***(First Revision)***1 SCOPE**

This specification covers the requirements of filter and filter chamber used in the blood transfusion apparatus.

2 REFERENCES

The standards given below contain provisions which, through reference in this text, constitute provisions of this standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this standard are encouraged to investigate the possibility of applying the most recent edition of these standards.

<i>IS No.</i>	<i>Title</i>
IS 1382 : 1981	Glossary of terms relating to glass and glassware (<i>first revision</i>)
IS 2303 (Part 1/ Sec 1) : 2021/ ISO 719 : 2020	Grading glass for alkalinity: Part 1 hydrolytic resistance of glass grains, Section 1 Determination and classification of hydrolytic resistance at 98 °C (<i>third revision</i>)
IS 3692 : 1975	Specification for rubber closures, pharmaceutical (<i>first revision</i>)

3 MATERIAL

3.1 The filter tube and chamber shall be made from clear, colorless, neutral glass (for definition see IS 1382). The glass shall pass the alkalinity test prescribed in IS 2303 (Part 1/Sec 1)/ISO 719 for Type I glass.

3.2 The filter shall be made from silk.

3.3 The bung shall be made from rubber; conforming to IS 3692.

4 SHAPE AND DIMENSIONS

The shape and dimensions shall be as per [Fig. 1](#).

5 WORKMANSHIP AND FINISH

5.1 The filter tube and chamber shall be well-annealed, free from bubbles and as far as possible, free from striae, stones and other visible defects (for definitions see IS 1382). The ends shall be smoothly rounded in the flame. It shall be capable of being easily cleaned. It shall pass the thermal shock test, dry heat test and autoclave test specified in [6.1](#), [6.2](#) and [6.3](#) respectively.

5.2 The filter shall have a filtering area of not less than 32 cm². The filter material shall be minimum of 80 percent as efficient as a sieve having a mesh with an average pore size of 0.212 mm square and a thread of 0.1 mm diameter (the reference filter material). The filter shall be disposed of after using it for one transfusion.

6 TESTS**6.1 Thermal Shock Test**

The filter tube and chamber shall be boiled in water for 30 min, then transferred to water at about 20 °C. The glass shall not develop any chipping or cracking.

6.2 Dry Heat Test

The filter tube and chamber shall be subjected to a dry heat test in a sterilizing oven at 180 °C ± 2 °C for 30 min. The glass shall not show deterioration in any way nor develop any crack or chipping.

6.3 Autoclave Test

The filter tube and chamber shall be autoclaved at a steam pressure of 1.4 kg/cm² for a period of 30 min. The glass shall not show deterioration in any way nor develop any crack or chipping.

7 MARKING

7.1 The filter and chamber shall be marked with the name of the manufacturer, his initials or trade-mark.

To access Indian Standards click on the link below:

https://www.services.bis.gov.in/php/BIS_2.0/bisconnect/knowyourstandards/Indian_standards/isdetails/

7.2 BIS Certification Marking


The product(s) conforming to the requirements of this standard may be certified as per the conformity assessment schemes under the provisions of the *Bureau of Indian Standards Act, 2016* and the Rules and Regulations framed thereunder, and the product(s) may be marked with the Standard Mark.

agreed to between the manufacturer and the purchaser. However, the recommended procedure is as follows:

‘Each set of filter tube and chamber shall be wrapped in suitable paper and packed in lots of 24 in suitable cartons’.

8 PACKING

The filter tube and chamber shall be packed as



All dimensions in millimetres (Nom).

FIG. 1 FILTER AND FILTER CHAMBER FOR BLOOD TRANSFUSION

ANNEX A

(Foreword)

COMMITTEE COMPOSITION

Hospital Equipment and Surgical Disposable Products Sectional Committee, MHD 12

<i>Organization</i>	<i>Representative(s)</i>
In Personal Capacity, (AIIMS Vijaypur - 184120)	LT GEN SUNIL KANT (Chairperson)
In Personal Capacity (Flat 315; Shelter Apt.; 15, Palm Grove Road; Victoria Layout - 60047)	SHRI KULVEEN SINGH BALI
3M India Limited, Bengaluru	DR PRABHA HEGDE MS KAVITHA KULKARNI (<i>Alternate</i>)
Asia Pacific Medical Technology Association (APACMed), Gurugram	SHRI R. ASHOK KUMAR SHRI PARVEEN JAIN (<i>Alternate</i>)
Association of Indian Medical Device Industry, New Delhi	SHRI RAVI ABRAHAM SHRI RAJIV NATH (<i>Alternate</i>)
B Braun Medical India Private Limited, New Delhi	SHRI VIVEK VEERBHAN MS ISHITA DHINGRA (<i>Alternate</i>)
B Medical Systems India Private Limited, New Delhi	SHRI KISHOR TUKARAM SHRI ANSHUMAN TULI (<i>Alternate</i>)
Boston Scientific India Private Limited, Gurugram	SHRI PRASHANTH PRABHAKAR SHRI DEV CHOPRA (<i>Alternate</i>)
Central Drugs Standard Control Organization, New Delhi	SHRI ASEEM SAHU MS SHYAMNI SASIDHARAN (<i>Alternate</i>)
ESIC Dental College and Hospital, New Delhi	SHRI NAGRAJ M. DR MANSI ATRI (<i>Alternate</i>)
Hindustan Syringes and Medical Devices Limited, Ballabgarh, Faridabad	SHRI PRAVEEN KUMAR SHARMA SHRI UPINDER VISHEN (<i>Alternate</i>)
Indian Rubber Gloves Manufacturers Association, New Delhi	SHRI MANMOHAN SINGH GULATI SHRI VIKAS ANAND (<i>Alternate</i>)
Johnson and Johnson Private Limited, Mumbai	SHRI HEMANT SONAWANE
Kalam Institute of Health Technology, Vishakhapatnam	SHRI AMIT SHARMA SHRI MOHAN RAGUL (<i>Alternate</i>)
Kanam Latex India Private Limited, Kottayam	SHRI ABRAHAM C. JACOB SHRI DONALD S. K. (<i>Alternate</i>)
Microtrol Sterilization Services Private Limited, Mumbai	SHRI BANSIDHAR S DHURANDHAR SHRI MANOJ MISHRA (<i>Alternate</i>)
National Institute of Health and Family Welfare, New Delhi	SHRI HITESH KUMAR SHRI SHIVLEY SAGEER (<i>Alternate</i>)

<i>Organization</i>	<i>Representative(s)</i>
Post Graduate Institute of Medical Education and Research, Chandigarh	DR NAVNEET DHALIWAL DR SHWETA TALATI (<i>Alternate I</i>) SHRI SANJEEV SHARMA (<i>Alternate II</i>)
Shriram Institute for Industrial Research, New Delhi	DR SANJAY RAJPUT MS MANISH RAWAT (<i>Alternate</i>)
Terumo Penpol Private Limited, Thiruvananthapuram	SHRI MANOJ A. SHRI V. M. SHAJAHAN (<i>Alternate</i>)
BIS Directorate General	SHRI A. R. UNNIKRIISHNAN SCIENTIST 'G' AND HEAD (MEDICAL EQUIPMENT AND HOSPITAL PLANNING) (<i>Ex-officio</i>)

Member Secretary

MS. UROOSA WARSI,
SCIENTIST 'C'/DEPUTY DIRECTOR
(MEDICAL EQUIPMENT AND HOSPITAL PLANNING), BIS

Bureau of Indian Standards

BIS is a statutory institution established under the *Bureau of Indian Standards Act, 2016* to promote harmonious development of the activities of standardization, marking and quality certification of goods and attending to connected matters in the country.

Copyright

BIS has the copyright of all its publications. No part of these publications may be reproduced in any form without the prior permission in writing of BIS. This does not preclude the free use, in the course of implementing the standard, of necessary details, such as symbols and sizes, type or grade designations. Enquiries relating to copyright be addressed to the Head (Publication & Sales), BIS.

Review of Indian Standards

Amendments are issued to standards as the need arises on the basis of comments. Standards are also reviewed periodically; a standard along with amendments is reaffirmed when such review indicates that no changes are needed; if the review indicates that changes are needed, it is taken up for revision. Users of Indian Standards should ascertain that they are in possession of the latest amendments or edition by referring to the website-www.bis.gov.in or www.standardsbis.in.

This Indian Standard has been developed from Doc No.: MHD 12 (22897).

Amendments Issued Since Publication

Amend No.	Date of Issue	Text Affected

BUREAU OF INDIAN STANDARDS

Headquarters:

Manak Bhavan, 9 Bahadur Shah Zafar Marg, New Delhi 110002

Telephones: 2323 0131, 2323 3375, 2323 9402

Website: www.bis.gov.in

Regional Offices:

	Telephones
Central : 601/A, Konnectus Tower -1, 6 th Floor, DMRC Building, Bhavbhuti Marg, New Delhi 110002	{ 2323 7617
Eastern : 8 th Floor, Plot No 7/7 & 7/8, CP Block, Sector V, Salt Lake, Kolkata, West Bengal 700091	{ 2367 0012 2320 9474
Northern : Plot No. 4-A, Sector 27-B, Madhya Marg, Chandigarh 160019	{ 265 9930
Southern : C.I.T. Campus, IV Cross Road, Taramani, Chennai 600113	{ 2254 1442 2254 1216
Western : 5 th Floor/MTNL CETTM, Technology Street, Hiranandani Gardens, Powai Mumbai 400076	{ 25700030 25702715

Branches : AHMEDABAD, BENGALURU, BHOPAL, BHUBANESHWAR, CHANDIGARH, CHENNAI, COIMBATORE, DEHRADUN, DELHI, FARIDABAD, GHAZIABAD, GUWAHATI, HARYANA (CHANDIGARH), HUBLI, HYDERABAD, JAIPUR, JAMMU, JAMSHEDPUR, KOCHI, KOLKATA, LUCKNOW, MADURAI, MUMBAI, NAGPUR, NOIDA, PARWANOO, PATNA, PUNE, RAIPUR, RAJKOT, SURAT, VIJAYAWADA.