

**केन्द्रीय मुहर विभाग – III**

हमारा संदर्भ- केन्द्रीय मुहर विभाग-III/16: IS 10258

31 May 2021

विषय: पनुरीक्षित प्रोडक्ट मैनुअल – एसआईटी एवं ग्रुपिंग गाइडलाइंस के साथ

सभी शाखा कार्यालय से आग्रह है कि उपरोक्त विषय से संबंधित संलग्न परिपत्र का अवलोकन करें।

(जे एस कविन कीर्थी विनायगम)  
वैज्ञानिक 'सी' (सी एम डी-III)

**प्रमुख (सी एम डी-III)**

सभी क्षेत्रीय/शाखा कार्यालय

**CENTRAL MARKS DEPARTMENT-III**

Our Ref: CMD-III/16: IS 10258

31 May 2021

Subject: Revised Product Manual incorporating SIT and Grouping Guidelines for IS 10258: 2002

BOs may kindly see the attached circular on the subject mentioned above.

(JS Kavin Keerthy Vinayagam)  
Sc-C (CMD III)

**Head (CMD - III)**

Circulated to: All ROs/BOs

**CENTRAL MARKS DEPARTMENT-III**

Our Ref: **CMD-III/Product Manual/IS 10258**

**31 May 2021**

**Subject: Revised Product Manual incorporating SIT and Grouping Guidelines for IS 10258: 2002**

This has reference to the subject mentioned above.

The approved Product Manual (PM/IS 10258/2/May 2021) for **IS 10258: 2002 - Sterile Hypodermic Syringes for single use** has been hosted on the BIS website under

**“Conformity Assessment → Product Certification → Product Specific Guidelines → Product Manual”.**

In this revision the frequency of tests at Cl. 12.1 (Piston detachment test) and 14.2 (Freedom from air and liquid leakage past piston) has been modified. The SIT, which has been incorporated in the Product Manual, shall be implemented **with immediate effect.**

**(JS Kavin Keerthy Vinayagam)**  
**Sc-C (CMD III)**

**Head (CMD III)**

**DDG (Certification)**

**Circulated to: All ROs/BOs**