

244

SYNOPSIS

Doc: MHD 15 (12819) IS 13450 (Part 2/Sec 34 : 2019/ IEC 60601-2-34 : 2011 Medical Electrical Equipment Part 2 Particular Requirements for the Basic Safety and Essential Performance Section 34 Invasive blood pressure monitoring equipment

a) Scope:

This particular standard applies to Basic Safety and Essential Performance of Invasive blood Pressure Monitoring equipment, hereinafter referred to as ME Equipment.

b) Salient features of content:

This particular standard does not apply to catheter tubing, catheter needles, Luer locks, taps and tap tables that connect to the Dome.

This particular standard does not apply to Non-Invasive Blood Pressure Monitoring Equipment.

c) Type/grades/classes, if any covered in the standard: Nil.

d) Disclaimer (to be automatically provided by the program/ software)
