

**BUREAU OF INDIAN STANDARDS**

**DRAFT FOR COMMENTS ONLY**

*(Not to be reproduced without permission of BIS or used as an Indian Standard)*

भारतीय मानक मसौदा  
**चिकित्सीय उपयोग के लिए इन्फ्यूजन उपकरण**  
**भाग 8 दबाव इन्फ्यूजन उपकरण के साथ एकल उपयोग के लिए इन्फ्यूजन सेट**

*Draft Indian Standard*

**Infusion Equipment for Medical Use**

**Part 8 Infusion Sets for Single Use with Pressure Infusion Apparatus**

ICS 11.040.20

Hospital Equipment and Surgical Disposable  
Products Sectional Committee, MHD 12

Last date for comments: **06 June 2024**

**NATIONAL FOREWORD**

*(Adoption clause will be added later)*

The text of ISO Standard has been approved as suitable for publication as an Indian Standard without deviations. Certain conventions are however not identical to those used in Indian Standards. Attention is particularly drawn to the following:

- Wherever the words 'International Standard' appear referring to this standard, they should be read as 'Indian Standard'
- Comma (,) has been used as a decimal marker while in Indian Standards, the current practice is to use a point (.) as the decimal marker.

In this adopted standard, reference appears to certain International Standards for which Indian Standards also exist. The corresponding Indian Standards which are to be substituted in their respective places are listed below along with their degree of equivalence for the editions indicated:

*International Standard*

*Corresponding Indian Standard*

*Degree of  
Equivalence*

|  |   |           |
|--|---|-----------|
| ISO 7000, Graphical symbols for use on equipment — Registered symbols  | IS 16450 : 2023/ISO 7000 : 2019, Graphical Symbols for Use on Equipment — Registered Symbols ( <i>First Revision</i> )  | Identical |
| ISO 8536-4, Infusion equipment for medical use — Part 4: Infusion sets for single use, gravity feed                                  | IS/ISO 8536-4 : 2019, Infusion Equipment for Medical Use Part 4 Infusion Sets for Single Use, Gravity Feed  | Identical |
| ISO 15223-1, Medical devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General requirements | IS 18105 (Part 1) : 2023/ISO 15223-1 : 2021, Medical Devices — Symbols to be Used with Information to be Supplied by the Manufacturer Part 1 General Requirements ( <i>Third Revision</i> ) | Identical |

The technical committee has reviewed the provisions of the following International Standards referred in this adopted standard and has decided that they are acceptable for use in conjunction with this standard:

*International Standard      Title*

|             |  |
|-------------|--|
| ISO 80369-7 | Small-bore connectors for liquids and gases in healthcare applications — Part 7: Connectors for intravascular or hypodermic applications |
|-------------|--|

This standard includes modification of cross reference from ISO 594-2 to ISO 80369-7. Wherever reference to ISO 594-2 occurs in the text, ISO 80369-7 has to be substituted.

This standard also makes a reference to the BIS Certification Marking of the product. Details of which is given in National Annex A.

For the purpose of deciding whether a particular requirement of this standard is complied with, the final value, observed or calculated, expressing the result of a test shall be rounded off in accordance with IS 2: 2022 ‘Rules for rounding off numerical values (*second revision*)’.

**NATIONAL ANNEX A**  
(National Foreword)

**A-1 BIS CERTIFICATION MARKING**

The product(s) conforming to the requirements of this standard may be certified as per the conformity assessment schemes under the provisions of the *Bureau of Indian Standards Act, 2016* and the Rules and Regulations framed thereunder, and the product(s) may be marked with the Standard Mark.

**Scope**

This part of ISO 8536 gives users information on sterilized infusion sets for single use with pressure infusion apparatus up to a maximum of 200 kPa (2 bar).

In some countries, the national pharmacopoeia or other national regulations are legally binding and take precedence over this part of ISO 8536.

---

The technical content of the document has not been enclosed as it is identical with the corresponding ISO standard. For details, please refer to ISO 8536-8:2015 or kindly contact:

Head  
Medical Equipment and Hospital Planning Department  
Bureau of Indian Standards  
9 Bahadur Shah Zafar Marg  
New Delhi-110002  
Email: [mhd@bis.gov.in](mailto:mhd@bis.gov.in)  
[hmhd@bis.gov.in](mailto:hmhd@bis.gov.in)