

AMENDMENT NO. 1 DECEMBER 2024

TO

**IS 8086: 2024 REHABILITATION EQUIPMENT —
WHEELCHAIRS, FOLDING, JUNIOR SIZE —
SPECIFICATION**

(Second Revision)

(Page 2, clause 6.6) — Substitute the following for the existing clause:

6.6 The solid tyres used in castors shall be made of non-marking resilient rubber or Ethylene Vinyl Acetate (EVA) or Polyurethane (PU) or Thermoplastic Polyurethane (TPU) with a minimum hardness of 60 Shore 'A'

(Page 2, clause 6.7.1) — Substitute the following for the existing clause:

6.7.1 The rear wheels shall be of minimum 451 (20") size and shall have rims, tyres, tubes, nipples, spokes and washers conforming to the relevant Indian Standard. The solid tyres shall be made of non-marking resilient rubber or ethylene vinyl acetate (EVA) or Polyurethane (PU) or Thermoplastic Polyurethane (TPU) with a minimum hardness of 60 shore 'A'.

NOTE — In Type 1 wheelchairs, the rear wheels could be as small as the front wheels.

(Page 5, Table 1, Sl No. xiii) — Substitute the following for the existing dimensions:

xiii)	Wheel diameter	451 (20") <i>Min</i>
-------	----------------	----------------------