

AMENDMENT NO. 1 DECEMBER 2024
TO
IS 5143: 2024 ADJUSTABLE AXILLARY CRUTCHES —
SPECIFICATION
(Third Revision)

(Page 2, clause 5.2) — Substitute the following for the existing clause:

5.2 Side bars of the aluminium crutches shall be powder-coated or shall be anodized as per Grade AC 10 of IS 1868.

(Page 4, Figure 2) — Substitute 'FOUR HOLES, PITCH 25' for 'THREE HOLES, PITCH 35'.