

FORMAT FOR SENDING COMMENTS ON BIS DOCUMENTS

(Please use A4 size sheet of paper only and type within fields indicated. Comments on each clauses/sub-clauses/table/fig. etc.be started on a fresh box. Information in Column 4 should include reasons for the comments and suggestions for modified wording of the clauses when the existing text is found not acceptable. Adherence to this format facilitates Secretariat's work)

TITLE: IS_9471 SPECIFICATION FOR MODULAR LOWER LIMB ORTHOTIC COMPONENTS, PART III: STIRRUP PLATES

NAME OF THE COMMENTATOR/ORGANIZATION: - Mr. VISHAL SHUKLA Sr. MANAGER (DD&QC), ALIMCO KANPUR

Sl. No.	Clause/Sub-clause/Para/table/fig. No. commented	Commentator/ Organization/ Abbreviation	Type of Comments (General/Editorial/ Technical)	Justification	Proposed change
1	Clause 3.1 Materials	ALIMCO	Technical	Due to thickness of Minimum 1.6 mm as mentioned in existing Clause 3.1of IS:9471 Part III, It restricts the use of sheet 1.6 mm as per IS:513 nad IS:1079 which is also in the mandatory list of BIS.	Plate - Mild steel sheet/strip cold rolled annealed Grade 0 of IS:513-1973 'Specification for cold rolled carbon steel sheets (second revision)' or hot rolled annealed Grade 0 of IS:1079-1973 'Specification for hot rolled carbon steel sheet and strip (third revision), Thickness 1.6mm.

